

**EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000**

**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST DAN RAISS AGENCY YAVAPAI Co. S.O.  
 DATE 4-28-08 TIME 1425  
 INTOXILYZER SERIAL # 80-001019 LOCATION YCSO-PRESCOTT JAIL

- (R) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
 OR  
 Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (R) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (R) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (R) 4. Air blank completed.
- (R) 5. Calibration check completed. Test results 0. 098 AC.
- (R) 6. Air blank completed.
- (R) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel A. Raiss #5-31 / OAS #33026

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number: 80-001019

04/28/2008

14:44:42

-----  
Standard Lot#: 730901I

Last Changed By: D. RAISS #3693

-----  
QAS: D. RAISS #3693

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	14:45:09
Cal Check	0.098	14:45:28
Air Blank	0.000	14:45:58

*David Q Rain #5-31/QAS #33026*  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST DAN RAISS AGENCY YAVAPAI CO. S.O.  
DATE 4-28-08 TIME 1425  
INTOXILYZER SERIAL # 80-001019 LOCATION YCSO - PRESCOTT JAIL

~~(P)~~ 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- ~~(P)~~ 1. Clock time check.  
~~(P)~~ 2. Date check.

OPERATIONAL TESTS

- ~~(P)~~ 1. Alcohol-free subject test result 0. 000 AC.  
~~(P)~~ 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
~~(P)~~ 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
~~(P)~~ 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS All OK.

SIGNATURE Dan Raiss #5-31/OAS #33026

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

04/28/2008

14:36:47

Standard Lot#: 730901I

Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693

YAVAPAI CSO

Operator: D. RAISS #3693

YAVAPAI CSO

Subject: SQAP,TEST,

DOB: 01/01/1901

Sex: M

Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:37:27
Diagnostic Test	Pass	14:37:54
Air Blank	0.000	14:38:22
0.100 Cal Check	0.097	14:38:41
Air Blank	0.000	14:39:10
Subject Test	0.114*	14:42:19
Air Blank	0.000	14:42:50
Air Blank	0.000	14:43:18
Subject Test	XXX**	14:43:42
Air Blank	0.000	14:44:15

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

04/28/2008

14:35:38

Standard Lot#: 730901I

Last Changed By: D. RAISS

#3693

QAS: D. RAISS

#3693

YAVAPAI CSO

Operator: D. RAISS

#3693

YAVAPAI CSO

Subject: SQAP,TEST,

DOB: 01/01/1901

Sex: M

Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	14:36:12
Air Blank	0.000	14:36:40

\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

04/28/2008

14:34:22

Standard Lot#: 730901I

Last Changed By: D. RAISS

#3693

QAS: D. RAISS

#3693

YAVAPAI CSO

Operator: D. RAISS

#3693

YAVAPAI CSO

Subject: SQAP,TEST,

DOB: 01/01/1901

Sex: M

Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	14:34:52
Air Blank	0.000	14:35:20

\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

04/28/2008

14:29:27

Standard Lot#: 730901I

Last Changed By: D. RAISS

#3693

QAS: D. RAISS

#3693

YAVAPAI CSO

Operator: D. RAISS

#3693

YAVAPAI CSO

Subject: SQAP,TEST,

DOB: 01/01/1901

Sex: M

Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:30:20
Diagnostic Test	Pass	14:30:47
Air Blank	0.000	14:31:15
0.100 Cal Check	0.097	14:31:33
Air Blank	0.000	14:32:03
Subject Test	0.000	14:32:24
Air Blank	0.000	14:32:51
Five Minute Wait	ABT*	
Air Blank	0.000	14:33:53

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**