

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST DAN RAISS AGENCY YAVAPAI Co. S.O.
DATE 8-20-08 TIME 2020
INTOXILYZER SERIAL # 80-001019 LOCATION YCSO-PRESCOTT JAIL

- (OR) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (OR) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (OR) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (OR) 4. Air blank completed.
- (OR) 5. Calibration check completed. Test results 0. 096 AC.
- (OR) 6. Air blank completed.
- (OR) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Manuel Q. Raiss #5-31 / OAS #33026

CAL CK PRIOR TO GAS CHANGE

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number: 80-001019

08/20/2008

20:38:40

Standard Lot#: 730901I

Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	20:39:07
Cal Check	0.096	20:39:26
Air Blank	0.000	20:39:55

Daniel J. Rain #5-31/QAS #33026

Operator's Signature

PRIOR

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST DAN RAISS AGENCY YAVAPAI Co. S.O.
DATE 8-20-08 TIME 2020
INTOXILYZER SERIAL # 80-001019 LOCATION YCSO-PRESCOTT JAIL

OK) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- OK) 1. Clock time check.
OK) 2. Date check.

OPERATIONAL TESTS

- OK) 1. Alcohol-free subject test result 0. 000 AC.
OK) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
OK) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
OK) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES X NO

COMMENTS ALL OK.

SIGNATURE Manuel Q Raim #5-31/ORS #33026

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number : 80-001019
08/20/2008

20:30:43

Standard Lot#: 730901I
Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693
YAVAPAI CSO

Operator: D. RAISS #3693
YAVAPAI CSO

Subject: SQAP,TEST,
DOB: 01/01/1901

Sex: M Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:31:26
Diagnostic Test	Pass	20:31:53
Air Blank	0.000	20:32:20
0.100 Cal Check	0.096	20:32:39
Air Blank	0.000	20:33:09
Subject Test	0.131*	20:36:17
Air Blank	0.000	20:36:49
Air Blank	0.000	20:37:16
Subject Test	XXX**	20:37:42
Air Blank	0.000	20:38:15

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number : 80-001019
08/20/2008

20:29:35

Standard Lot#: 730901I
Last Changed By: D. RAISS #3693
QAS: D. RAISS #3693
YAVAPAI CSO
Operator: D. RAISS #3693
YAVAPAI CSO

Subject: SQAP,TEST,
DOB: 01/01/1901
Sex: M Weight: 1
15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	20:29:56
Air Blank	0.000	20:30:24

*RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

08/20/2008

20:28:28

Standard Lot#: 730901I

Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693

YAVAPAI CSO

Operator: D. RAISS #3693

YAVAPAI CSO

Subject: SQAP, TEST,

DOB: 01/01/1901

Sex: M

Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	20:29:02
Air Blank	0.000	20:29:31

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number : 80-001019
08/20/2008

20:23:45

Standard Lot#: 730901I
Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693
YAVAPAI CSO

Operator: D. RAISS #3693
YAVAPAI CSO

Subject: SQAP,TEST,
DOB: 01/01/1901

Sex: M Weight: 1

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	20:25:13
Diagnostic Test	Pass	20:25:40
Air Blank	0.000	20:26:08
0.100 Cal Check	0.097	20:26:26
Air Blank	0.000	20:26:56
Subject Test	0.000	20:27:19
Air Blank	0.000	20:27:46
Five Minute Wait	ABT*	
Air Blank	0.000	20:28:21

*Sequence Aborted

Not a Successfully Completed Test Sequence