

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number: 80-001019

09/18/2009

21:44:48

Standard Lot#: 818201G

Last Changed By: D. RAISS #3693

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	21:45:16
Cal Check	0.097	21:45:34
Air Blank	0.000	21:46:04



Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 09-18-09 TIME 2155

INTOXILYZER SERIAL # 80-00109 LOCATION Prescott

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- DM 1. Clock time check.
- DM 2. Date check.

OPERATIONAL TESTS

- DM 1. Alcohol-free subject test result 0. 000 AC.
- DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- DM 4. Standard Calibration Check standard 0. 0.10 AC. Result 0. 0.16 AC.

Instrument is operating properly and accurately. YES DM NO _____

COMMENTS All OK

SIGNATURE Daniel McGinnis 343/QAS# 33426

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number : 80-001019
09/18/2009

21:55:21

Standard Lot#: 818201G
Last Changed By: D. RAISS #3693

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST

DOB: 09/09/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:56:30
Diagnostic Test	Pass	21:56:57
Air Blank	0.000	21:57:25
0.100 Cal Check	0.097	21:57:44
Air Blank	0.000	21:58:13
Subject Test	0.071*	22:01:22
Air Blank	0.000	22:01:53
Air Blank	0.000	22:02:21
Subject Test	XXX**	22:03:45
Air Blank	IPS***	22:03:53

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number: 80-001019

09/18/2009

22:05:52

Standard Lot#: 818201G


Last Changed By: D. RAISS #3693

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	AMB*	22:06:20
Air Blank	0.000	22:06:48

*Ambient Fail


Operator's Signature

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number : 80-001019

09/18/2009

21:50:54

Standard Lot#: 818201G

Last Changed By: D. RAISS

#3693

QAS: D. MCGINNIS

#8086

YAVAPAI CSO

Operator: D. MCGINNIS

#8086

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 09/09/1990

Sex: M

Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	21:52:09
Diagnostic Test	Pass	21:52:36
Air Blank	0.000	21:53:03
0.100 Cal Check	0.097	21:53:22
Air Blank	0.000	21:53:52
Subject Test	0.000	21:54:33
Air Blank	ABT*	21:54:42
Air Blank	RFI**	21:54:51

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YUSD

DATE 09-18-09 TIME 2200

INTOXILYZER SERIAL # 80-001029 LOCATION Prescott

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR

~~Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$~~

2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McGinnis

INTOXILYZER 8000

Location: YCSO PRESCOTT JAIL

Serial Number: 80-001019

09/18/2009

22:06:53

Standard Lot#: 818201G

Last Changed By: D. RAISS #3693

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:07:21
Cal Check	0.096	22:07:39
Air Blank	0.000	22:08:09


Operator's Signature