

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core version: 6105.48
06/20/2010 23:14:08

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

OAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Subject: SOAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:15:12
Diagnostic Test	Pass	23:15:45
Air Blank	0.000	23:16:16
0.100 Cal Check	0.097	23:16:37
Air Blank	0.000	23:17:07
Subject Test	0.020*	23:20:14
Air Blank	0.000	23:20:45
Air Blank	0.000	23:21:12
Subject Test	XXX**	23:22:05
Air Blank	IAS***	23:22:13

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***Improper Sample

Not a Successfully
Completed
Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core version: 6105.48
06/20/2010 23:10:05

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

OAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Subject: SOAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:11:05
Diagnostic Test	Pass	23:11:42
Air Blank	0.000	23:12:09
0.100 Cal Check	0.097	23:12:30
Air Blank	0.000	23:13:00
Subject Test	0.000	23:13:19
Air Blank	ABT*	23:13:30
Air Blank	RFI**	23:13:36

*Sequence Aborted
**RFI Detect

Not a Successfully
Completed
Test Sequence

INTOXILYZER 6000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core Version: 8105.48
06/20/2010 23:24:05

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

DAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:24:36
0.100 Cal Check	0.098	23:24:57
Air Blank	0.000	23:25:27

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core Version: 8105.48
06/20/2010 23:06:36

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

DAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:07:08
0.100 Cal Check	0.097	23:07:39
Air Blank	0.000	23:07:59

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core Version: 8105.48
06/20/2010 23:24:05

Standard Lot#: 712322
Last Changed By: D. RAJSS #3693

DAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:24:36
0.100 Cal Check	0.098	23:24:57
Air Blank	0.000	23:25:27

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000
Location: YCSO PRESCOTT JAIL
Serial Number: 80-001019
Core Version: 8105.48
06/20/2010 23:06:36

Standard Lot#: 712322
Last Changed By: D. RAJSS #3693

DAS: D. MCGINNIS #8086
YAUAPAI CSO

Operator: D. MCGINNIS #8086
YAUAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:07:06
0.100 Cal Check	0.097	23:07:25
Air Blank	0.000	23:07:58

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO
DATE 6-20-10 TIME 2310
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- DM 1. Clock time check.
DM 2. Date check.

OPERATIONAL TESTS

- DM 1. Alcohol-free subject test result 0. 000 AC.
DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
DM 4. Standard Calibration Check standard 0. 10 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES DM NO _____

COMMENTS All OK

SIGNATURE Daniel McGinnis S43/QAS#33426

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McInnis AGENCY YCSO
DATE 6-20-10 TIME 2324
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- DM* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- DM* 4. Air blank completed.
- DM* 5. Calibration check completed. Test results 0. 097 AC.
- DM* 6. Air blank completed.
- DM* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McInnis S43/QAS # 33426