

**COPY**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NSB AGENCY YCSO  
DATE 9/29/10 TIME 15:23:51  
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott Patrol

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .108 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.095 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NSB

INTOXILYZER 8000  
Location: YCSO PRESCOTT PATROL  
Serial Number: 80-001019  
Core Version: 8105.48  
09/29/2010 15:23:51

-----  
Standard Lot#: 712322  
Last Changed By: D. RAISS #3693  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	15:24:33
0.100 Cal Check	0.095	15:24:54
Air Blank	0.000	15:25:24

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000

Location: YCSO PRESCOTT PATROL

Serial Number: 80-001019

Core Version: 8105.48

09/29/2010

15:27:22

---

Standard Lot#: 712322

Last Changed By: D. RAISS

#3693

---

QAS: J. MURRAY

#N56

YAVAPAI YCSO

---

Operator: J. MURRAY

#N56

YAVAPAI YCSO

---

Subject: CALCHECK,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	15:29:14
Diagnostic Test	Pass	15:29:49
Air Blank	0.000	15:30:17
0.100 Cal Check	0.096	15:30:38
Air Blank	0.000	15:31:07
Subject Test	0.000	15:31:47
Air Blank	0.000	15:32:14
Wait		
Air Blank	0.000	15:37:14
Subject Test	0.000	15:37:36
Air Blank	0.000	15:38:04
0.100 Cal Check	0.098	15:38:24
Air Blank	0.000	15:38:54
Diagnostic Test	Pass	15:39:29

**Successfully Completed Test Sequence**

---