

COPY

**EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**

**ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST J. Murrby NSB AGENCY YC 50
DATE 10/17/10 TIME 8:32:27
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott Patrol

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.095 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murrby NSB

INTOXILYZER 8000

Location: YCSO PRESCOTT PATROL

Serial Number: 80-001019

Core Version: 8105.48

10/17/2010

08:32:27

Standard Lot#: 712322

Last Changed By: D. RAISS #3693

QAS: J. MURRAY #N56

YAVAPAI YCSO

Operator: J. MURRAY #N56

YAVAPAI YCSO

Test g/210L Time

Air Blank 0.000 08:33:09
0.100 Cal Check 0.095 08:33:29
Air Blank 0.000 08:33:59

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Murray NSL AGENCY YCSO
DATE 10/17/10 TIME 8:41:01
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott Patrol

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 095 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Gas Tank # 712322 300 PSI

SIGNATURE

J. Murray NSL

INTOXILYZER 8000
Location: YCSO PRESCOTT PATROL
Serial Number: 80-001019
Core Version: 8105.48
10/17/2010 08:41:01

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Subject: TEST,SQAP
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:42:58
Diagnostic Test	Pass	08:43:33
Air Blank	0.000	08:44:00
0.100 Cal Check	0.096	08:44:21
Air Blank	0.000	08:44:51
Subject Test	0.000	08:45:11
Air Blank	0.000	08:45:38
Wait	ABT*	
Air Blank	RFI**	08:46:33

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT PATROL
Serial Number: 80-001019
Core Version: 8105.48
10/17/2010 08:48:06

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	08:49:08
Air Blank	0.000	08:49:36

*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT PATROL
Serial Number: 80-001019
Core Version: 8105.48
10/17/2010 08:49:45

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:51:08
Diagnostic Test	Pass	08:51:43
Air Blank	0.000	08:52:11
0.100 Cal Check	0.095	08:52:32
Air Blank	0.000	08:53:01
Subject Test	0.081*	08:56:08
Air Blank	0.000	08:56:40
Air Blank	0.000	08:57:07
Subject Test	XXX**	08:57:47
Air Blank	0.000	08:58:21

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NSB AGENCY YCSO
DATE 10/17/10 TIME 8 59 17
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott Patrol

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.095 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NSB

INTOXILYZER 8000
Location: YCSO PRESCOTT PATROL
Serial Number: 80-001019
Core Version: 8105.48
10/17/2010 08:59:17

Standard Lot#: 712322
Last Changed By: D. RAISS #3693

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Test	g/210L	Time
Air Blank	0.000	08:59:57
0.100 Cal Check	0.095	09:00:18
Air Blank	0.000	09:00:47

Preventative Maintenance
Performed -- Timer Reset