

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NS6/5597 AGENCY Yavapai County Sheriff's
DATE 1/27/2013 TIME 08 28 09
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott 103

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 096 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000
Location: YCSO PRESCOTT
Serial Number: 80-001019
Core Version: 8105.48
01/27/2013 08:28:09

Standard Lot#: 06612100A1
Last Changed By: L. MOODY #7486

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Test	g/210L	Time
Air Blank	0.000	08:28:43
0.100 Cal Check	0.096	08:29:04
Air Blank	0.000	08:29:34

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. MULLAY NS6/5597 AGENCY Yavapai County Sheriff's Office
DATE 1/27/2013 TIME 08 35 26
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott 103

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS SNA# 06622180A1 exp-5/1/2014
+300 PSI

SIGNATURE J. MULLAY NS6/5597

INTOXILYZER 8000
Location: YCSO PRESCOTT
Serial Number: 80-001019
Core Version: 8105.48
01/27/2013

08:35:26

Standard Lot#: 06612100A1
Last Changed By: L. MOODY #7486

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:36:22
Diagnostic Test	Pass	08:36:57
Air Blank	0.000	08:37:24
0.100 Cal Check	0.095	08:37:45
Air Blank	0.000	08:38:14
Subject Test	0.000	08:38:49
Air Blank	0.000	08:39:16
Wait	ABT*	
Air Blank	IPS**	08:39:31

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO PRESCOTT

Serial Number: 80-001019

Core Version: 8105.48

01/27/2013

08:39:38

Standard Lot#: 06612100A1

Last Changed By: L. MOODY

#7486

QAS: J. MURRAY

#N56

YAVAPAI YCSO

Operator: J. MURRAY

#N56

YAVAPAI YCSO

Subject: TEST,SQAP

DOB: 10/20/1960

Sex: F

Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:40:19
Air Blank	IPS**	08:40:26

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO PRESCOTT
Serial Number: 80-001019
Core Version: 8105.48
01/27/2013

08:40:32

Standard Lot#: 06612100A1
Last Changed By: L. MOODY #7486

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Subject: SQAP,TEST
DOB: 10/20/1960
Sex: F Weight: 115

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:41:29
Diagnostic Test	Pass	08:42:04
Air Blank	0.000	08:42:32
0.100 Cal Check	0.096	08:42:52
Air Blank	0.000	08:43:22
Subject Test	0.072*	08:46:29
Air Blank	0.000	08:47:01
Air Blank	0.000	08:47:28
Subject Test	XXX**	08:48:01
Air Blank	0.000	08:48:34

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NSB/5597 AGENCY Yavapai County Sheriff's
DATE 1/27/2013 TIME 08 48 48
INTOXILYZER SERIAL # 82-001019 LOCATION Prescott 103

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 094 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NSB/5597

INTOXILYZER 8000
Location: YCSO PRESCOTT
Serial Number: 80-001019
Core Version: 8105.48
01/27/2013 08:48:48

Standard Lot#: 06612100A1
Last Changed By: L. MOODY #7486

QAS: J. MURRAY #N56
YAVAPAI YCSO

Operator: J. MURRAY #N56
YAVAPAI YCSO

Test	g/210L	Time
Air Blank	0.000	08:49:27
0.100 Cal Check	0.094	08:49:48
Air Blank	0.000	08:50:17

Preventative Maintenance
Performed -- Timer Reset