

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai Co. S.O.  
DATE 2/27/2013 TIME 7 29 57  
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott 103

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .106 AC.  
     Pour a standard alcohol concentration solution OR AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0. 095 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

300 PSI

006612100A1  
.5/1/2014

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
02/27/2013 07:29:57

-----  
Standard Lot#: 06612100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	07:31:01
0.100 Cal Check	0.095	07:31:22
Air Blank	0.000	07:31:54

Preventative Maintenance  
Performed -- Timer Reset