

# COPY

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. MURRAY NSB 15597 AGENCY Yavapai County Sher, H's Office  
DATE 4/24/13 TIME 7:35:22  
INTOXILYZER SERIAL # 80-601019 LOCATION Prescott 103

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.095 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. MURRAY NSB 15597

INTOXILYZER 8000

Location: YCSO PRESCOTT

Serial Number: 80-001019

Core Version: 8105.48

04/21/2013

07:35:22

-----  
Standard Lot#: 32012100A1

Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56

YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56

YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	07:35:57
0.100 Cal Check	0.095	07:36:18
Air Blank	0.000	07:36:50

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Murray NSB/5597 AGENCY Yavapai County Sheriff's Office  
DATE 4/24/13 TIME 07:37:01  
INTOXILYZER SERIAL # 86-001019 LOCATION Prescott 103

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

7/1/2015 32012 100A1  
900PSI

SIGNATURE J. Murray NSB/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
04/21/2013 07:37:01

---

Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:37:57
Diagnostic Test	Pass	07:38:32
Air Blank	0.000	07:38:59
0.100 Cal Check	0.094	07:39:20
Air Blank	0.000	07:39:52
Subject Test	0.000	07:40:29
Air Blank	0.000	07:40:57
Wait	ABT*	
Air Blank	RFI**	07:41:28

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
04/21/2013

07:41:35

---

Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	07:42:31
Air Blank	RFI**	07:42:39

\*Improper Sample  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
04/21/2013

07:42:52

---

Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:43:55
Diagnostic Test	Pass	07:44:30
Air Blank	0.000	07:44:57
0.100 Cal Check	0.095	07:45:18
Air Blank	0.000	07:45:50
Subject Test	0.075*	07:48:58
Air Blank	0.000	07:49:29
Air Blank	0.000	07:49:56
Subject Test	XXX**	07:50:29
Air Blank	0.000	07:51:02

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. MURRAY NS6/5597 AGENCY Yavapai County Sheriff's Office  
DATE 4/21/13 TIME 07:57:59  
INTOXILYZER SERIAL # 80-001019 LOCATION PASCOFF 103

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 094 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
04/21/2013 07:51:59

-----  
Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	07:52:33
0.100 Cal Check	0.094	07:52:54
Air Blank	0.000	07:53:26

Preventative Maintenance  
Performed -- Timer Reset