

Copy

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Ashby 815/2271 AGENCY YCSO  
DATE 5.23.13 TIME 1843  
INTOXILYZER SERIAL # 80.001019 LOCATION Prescott 103

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.094 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Dawn Ashby #2271 / QAS #33400

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
05/23/2013 18:43:00

-----  
Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56  
-----

QAS: D. ASHBY #S62  
YAVAPAI CSO  
-----

Operator: D. ASHBY #S62  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:43:36
0.100 Cal Check	0.094	18:43:57
Air Blank	0.000	18:44:28

Preventative Maintenance  
Performed -- Timer Reset