

COPY

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST M. V. JAY AGENCY YES O  
DATE 7/21/13 TIME 12 55 08  
INTOXILYZER SERIAL # 80-001019 LOCATION Pinecroft 103

- 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
     OR  
     Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4. Air blank completed.
- 5. Calibration check completed. Test results 0.093 AC.
- 6. Air blank completed.
- 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. M. V. JAY NS2/5597

INTOXILYZER 8000

Location: YCSO PRESCOTT

Serial Number: 80-001019

Core Version: 8105.48

07/21/2013

12:55:08

-----  
Standard Lot#: 32012100A1

Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56

YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56

YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	12:55:54
0.100 Cal Check	0.093	12:56:15
Air Blank	0.000	12:56:45

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Murphy AGENCY YCSO  
DATE 7/2/13 TIME 12:58:18  
INTOXILYZER SERIAL # 80-00 1019 LOCATION Pilsbitt 103

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 091 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

32012100A1 1/1/2015 500+ PSI

SIGNATURE J. Murphy NSB/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
07/21/2013 12:58:18

---

Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:59:15
Diagnostic Test	Pass	12:59:50
Air Blank	0.000	13:00:18
0.100 Cal Check	0.092	13:00:38
Air Blank	0.000	13:01:08
Subject Test	0.000	13:01:37
Air Blank	0.000	13:02:05
Wait	ABT*	
Air Blank	RFI**	13:02:24

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO PRESCOTT

Serial Number: 80-001019

Core Version: 8105.48

07/21/2013

13:02:34

---

Standard Lot#: 32012100A1

Last Changed By: J. MURRAY

#N56

---

QAS: J. MURRAY

#N56

YAVAPAI YCSO

---

Operator: J. MURRAY

#N56

YAVAPAI YCSO

---

Subject: SQAP,TEST

DOB: 10/20/1960

Sex: F

Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	13:03:16
Air Blank	RFI**	13:03:23

\*Improper Sample

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
07/21/2013

13:03:28

---

Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56

---

QAS: J. MURRAY #N56  
YAVAPAI YCSO

---

Operator: J. MURRAY #N56  
YAVAPAI YCSO

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:04:22
Diagnostic Test	Pass	13:04:57
Air Blank	0.000	13:05:24
0.100 Cal Check	0.092	13:05:45
Air Blank	0.000	13:06:14
Subject Test	0.064*	13:09:22
Air Blank	0.000	13:09:53
Air Blank	0.000	13:10:21
Subject Test	XXX**	13:11:35
Air Blank	0.000	13:12:08

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Murphy AGENCY YCSO  
DATE 7/21/13 TIME 13:12:30  
INTOXILYZER SERIAL # P0001019 LOCATION Prescott 103

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. 091 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murphy NSB/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
07/21/2013 13:12:30

-----  
Standard Lot#: 32012100A1  
Last Changed By: J. MURRAY #N56  
-----

QAS: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Operator: J. MURRAY #N56  
YAVAPAI YCSO  
-----

Test	g/210L	Time
Air Blank	0.000	13:13:07
0.100 Cal Check	0.091	13:13:28
Air Blank	0.000	13:13:58

Preventative Maintenance  
Performed -- Timer Reset