

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray AGENCY Yavapai County Sheriff's Office
DATE Sept. 21, 2013 TIME 07:50:27
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott Intox Room

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (+) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(+) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(+) 4. Air blank completed.
(+) 5. Calibration check completed. Test results 0.093 AC.
(+) 6. Air blank completed.
(+) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000

Location: YCSO PRESCOTT

Serial Number: 80-001019

Core Version: 8105.48

09/21/2013

07:50:27

Standard Lot#: 32012100A1

Last Changed By: J. MURRAY #N56

QAS: J. MURRAY 5597 #N56

YAVAPAI CSO

Operator: J. MURRAY 5597 #N56

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	07:51:47
0.100 Cal Check	0.093	07:52:08
Air Blank	0.000	07:52:38

Preventative Maintenance
Performed -- Timer Reset