

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

## STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NSB/5597 AGENCY Yavapai County Sheriff's Office  
DATE 9/28/2014 TIME 06:56:43  
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott

- (+) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (+) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (+) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (+) 4. Air blank completed.
- (+) 5. Calibration check completed. Test results 0.095 AC.
- (+) 6. Air blank completed.
- (+) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NSB/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
09/28/2014 06:56:43

-----  
Standard Lot#: 15013100A2  
Last Changed By: J. SLOWINSKI #6101  
-----

QAS: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS  
-----

Operator: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS  
-----

Test	g/210L	Time
Air Blank	0.000	06:57:27
0.100 Cal Check	0.095	06:57:47
Air Blank	0.000	06:58:17

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST J. Mully NSB/5597 AGENCY Yavapai County Sheriff's Office  
DATE 9/28/2014 TIME 06:59:50  
INTOXILYZER SERIAL # 80-001619 LOCATION Prescott

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

150/3100A2 6/1/2015  
-400 PSI

SIGNATURE J. Mully NSB/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
09/28/2014 06:59:50

---

Standard Lot#: 15013100A2  
Last Changed By: J. SLOWINSKI #6101

---

QAS: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Operator: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:00:52
Diagnostic Test	Pass	07:01:27
Air Blank	0.000	07:01:54
0.100 Cal Check	0.095	07:02:15
Air Blank	0.000	07:02:45
Subject Test	0.000	07:03:26
Air Blank	0.000	07:03:53
Wait	ABT*	
Air Blank	IPS**	07:04:08

\*Sequence Aborted

\*\*Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
09/28/2014 07:04:11

---

Standard Lot#: 15013100A2  
Last Changed By: J. SLOWINSKI #6101

---

QAS: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Operator: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	RFI*	07:04:52
Air Blank	IPS**	07:05:04

\*RFI Detect

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
09/28/2014

07:05:41

---

Standard Lot#: 15013100A2  
Last Changed By: J. SLOWINSKI #6101

---

QAS: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Operator: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS

---

Subject: SQAP,TEST  
DOB: 10/20/1960  
Sex: F Weight: 115

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	07:06:42
Diagnostic Test	Pass	07:07:17
Air Blank	0.000	07:07:44
0.100 Cal Check	0.095	07:08:05
Air Blank	0.000	07:08:34
Subject Test	0.253*	07:11:42
Air Blank	0.000	07:12:15
Air Blank	0.000	07:12:42
Subject Test	XXX**	07:13:16
Air Blank	0.000	07:13:49

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST J. Murray NS6/5597 AGENCY Yavapai County Sheriff's Office  
DATE 9/28/2014 TIME 07:14:03  
INTOXILYZER SERIAL # 80-001019 LOCATION Prescott

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution        AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0.094 AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE J. Murray NS6/5597

INTOXILYZER 8000  
Location: YCSO PRESCOTT  
Serial Number: 80-001019  
Core Version: 8105.48  
09/28/2014 07:14:03

-----  
Standard Lot#: 15013100A2  
Last Changed By: J. SLOWINSKI #6101  
-----

QAS: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS  
-----

Operator: J. MURRAY5597 #N56  
YAVAPAI SHERIFFS  
-----

Test	g/210L	Time
Air Blank	0.000	07:14:42
0.100 Cal Check	0.094	07:15:03
Air Blank	0.000	07:15:32

Preventative Maintenance  
Performed -- Timer Reset