

CA

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST DAN RAISS AGENCY YAVAPAI Co. S.O.
DATE 9-15-08 TIME 0105
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAYER

- (OK) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (OK) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (OK) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (OK) 4. Air blank completed.
- (OK) 5. Calibration check completed. Test results 0. 097 AC.
- (OK) 6. Air blank completed.
- (OK) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel Q. Raiss #5-31/OAS#33026

INTOXILYZER 8000

Location: YCSO-MAYER

Serial Number: 80-001095

09/15/2008

01:07:17


Standard Lot#: 733201G

Last Changed By: D. RAISS #3693

QAS: D. RAISS #3693

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	01:07:44
Cal Check	0.097	01:08:03
Air Blank	0.000	01:08:33


Operator's Signature