

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO
DATE 6-12-10 TIME 2315
INTOXILYZER SERIAL # 80-00095 LOCATION Mayer

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 110 AC.
OR
Pour a ~~standard alcohol concentration solution~~ AC, into a ~~clean dry simulator~~ and assemble the simulator. ~~Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
DM 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
DM 4. Air blank completed.
DM 5. Calibration check completed. Test results 0. 097 AC.
DM 6. Air blank completed.
DM 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Daniel McGinnis 543/QAS 33426

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
06/12/2010 23:15:54

Standard Lot#: 901401I
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:16:25
0.100 Cal Check	0.097	23:16:46
Air Blank	0.000	23:17:16

Preventative Maintenance
Performed -- Timer Reset