

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS  
DATE 11/20/10 TIME 1251  
INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

- (RJB) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (RJB) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (RJB) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (RJB) 4. Air blank completed.
- (RJB) 5. Calibration check completed. Test results 0.099 AC.
- (RJB) 6. Air blank completed.
- (RJB) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III # 6396

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

11/20/2010

12:51:42

-----  
Standard Lot#: 590854

Last Changed By: R. BYRNE, III #6386

-----  
QAS: R. BYRNE, III #6386

AZDPS

-----  
Operator: R. BYRNE, III #6386

AZDPS

-----  
Test                      g/210L                      Time  
-----  
Air Blank                      0.000                      12:52:18  
0.100 Cal Check                0.099                      12:52:39  
Air Blank                      0.000                      12:53:08

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY 'A2 DPS

DATE 11/20/10 TIME 1254

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D.

**HB**) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- HB**) 1. Clock time check.  
( ) 2. Date check.

OPERATIONAL TESTS

- HB** 1. Alcohol-free subject test result 0. 000 AC.  
**HB** 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
**HB** 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
**HB** 4. Standard Calibration Check standard 0. 100 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES **HB** NO \_\_\_\_\_

COMMENTS

LOT #590854 EXP 03/05/11

SIGNATURE *Richard J. Byrne III* #6386

INTOXILYZER 8000  
Location: WICKENBURG PD  
Serial Number: 80-000929  
Core Version: 8105.48  
11/20/2010

12:54:58

---

Standard Lot#: 590854  
Last Changed By: R. BYRNE, III #6386

---

QAS: R. BYRNE, III #6386  
AZDPS

---

Operator: R. BYRNE, III #6386  
AZDPS

---

Subject: SQAP,TEST  
DOB: 08/08/1968  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	12:56:51
Diagnostic Test	Pass	12:57:26
Air Blank	0.000	12:57:53
0.100 Cal Check	0.100	12:58:14
Air Blank	0.000	12:58:44
Subject Test	0.000	12:59:09
Air Blank	0.000	12:59:37
Wait	ABT*	
Air Blank	0.000	13:00:12

\*Sequence Aborted

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: WICKENBURG PD  
Serial Number: 80-000929  
Core Version: 8105.48  
11/20/2010

13:00:27

---

Standard Lot#: 590854  
Last Changed By: R. BYRNE, III #6386

---

QAS: R. BYRNE, III #6386  
AZDPS

---

Operator: R. BYRNE, III #6386  
AZDPS

---

Subject: SQAP, TEST  
DOB: 08/08/1968  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	IPS*	13:01:06
Air Blank	RFI**	13:01:13

\*Improper Sample  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: WICKENBURG PD  
Serial Number: 80-000929  
Core Version: 8105.48  
11/20/2010

13:01:41

---

Standard Lot#: 590854  
Last Changed By: R. BYRNE, III #6386

---

OAS: R. BYRNE, III #6386  
AZDPS

---

Operator: R. BYRNE, III #6386  
AZDPS

---

Subject: SQAP,TEST  
DOB: 08/08/1968  
Sex: M Weight: 180

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:02:44
Diagnostic Test	Pass	13:03:19
Air Blank	0.000	13:03:46
0.100 Cal Check	0.100	13:04:07
Air Blank	0.000	13:04:37
Subject Test	0.142*	13:07:44
Air Blank	0.000	13:08:16
Air Blank	0.000	13:08:43
Subject Test	XXX**	13:09:09
Air Blank	0.000	13:09:43

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RICHARD J. BYRNE III AGENCY AZ DPS

DATE 11/20/10 TIME 1310

INTOXILYZER SERIAL # 80-000929 DPS LOCATION WICKENBURG P.D

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution N/A AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. \_\_\_\_\_ AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Richard J. Byrne III #6386

INTOXILYZER 8000

Location: WICKENBURG PD

Serial Number: 80-000929

Core Version: 8105.48

11/20/2010

13:10:44

-----  
Standard Lot#: 590854

Last Changed By: R. BYRNE, III #6386  
-----

QAS: R. BYRNE, III #6386

AZDPS  
-----

Operator: R. BYRNE, III #6386

AZDPS  
-----

Test	g/210L	Time
Air Blank	0.000	13:11:15
0.100 Cal Check	0.100	13:11:36
Air Blank	0.000	13:12:06

Preventative Maintenance  
Performed -- Timer Reset