

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 12-16-10 TIME 7047

INTOXILYZER SERIAL # 80-001095 LOCATION Mayer

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 1.10 AC  
OR  
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- DM* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- DM* 4. Air blank completed.
- DM* 5. Calibration check completed. Test results 0. 097 AC.
- DM* 6. Air blank completed.
- DM* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Daniel McGinnis S43/QAS# 33424

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
12/16/2010 20:47:54

-----  
Standard Lot#: 901401I  
Last Changed By: D. MCGINNIS #8086  
-----

QAS: D. MCGINNIS #8086  
YAVAPAI CSO  
-----

Operator: D. MCGINNIS #8086  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:48:29
0.100 Cal Check	0.097	20:48:49
Air Blank	0.000	20:49:19

Preventative Maintenance  
Performed -- Timer Reset