

EXHIBIT G-2  
FIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104.5)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 08-02-11 TIME 1739

INTOXILYZER SERIAL # 8-001095 LOCATION Mayer

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 10 AC  
OR  
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. ~~Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 096 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Daniel McGinnis 343/016#33426

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
08/02/2011 17:39:44

-----  
Standard Lot#: 20910100A1  
Last Changed By: D. MCGINNIS #8086  
-----

QAS: D. MCGINNIS #8086  
YAVAPAI CSO  
-----

Operator: D. MCGINNIS #8086  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	17:40:14
0.100 Cal Check	0.096	17:40:35
Air Blank	0.000	17:41:04

Preventative Maintenance  
Performed -- Timer Reset