

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 01-13-2012 TIME 1955

INTOXILYZER SERIAL # 80-001095 LOCATION MAYER

- (  ) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution of 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (  ) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (  ) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (  ) 4. Air blank completed.
- (  ) 5. Calibration check completed. Test results 0. 097 AC.
- (  ) 6. Air blank completed.
- (  ) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LMoody 1748

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
01/13/2012 19:55:33

-----  
Standard Lot#: 20910100A1  
Last Changed By: J. JAEGER #3566  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	19:59:44
0.100 Cal Check	0.097	20:00:05
Air Blank	0.000	20:00:34

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 01-13-2012 TIME 2002

INTOXILYZER SERIAL # 80-001095 LOCATION MAYER

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.  
(  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.  
(  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(  ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

TANK @ 800psi

SIGNATURE Logan Moody

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
01/13/2012 20:02:42

---

Standard Lot#: 20910100A1  
Last Changed By: J. JAEGER #3566

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 190

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:03:40
Diagnostic Test	Pass	20:04:14
Air Blank	0.000	20:04:41
0.100 Cal Check	0.097	20:05:02
Air Blank	0.000	20:05:31
Subject Test	0.000	20:05:56
Air Blank	0.000	20:06:23
Wait	ABT*	
Air Blank	RFI**	20:06:51

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
01/13/2012

20:06:56

---

Standard Lot#: 20910100A1  
Last Changed By: J. JAEGER #3566

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 190

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:07:45
Diagnostic Test	Pass	20:08:19
Air Blank	0.000	20:08:47
0.100 Cal Check	0.096	20:09:07
Air Blank	0.000	20:09:37
Subject Test	0.080*	20:12:44
Air Blank	0.000	20:13:16
Air Blank	0.000	20:13:43
Subject Test	XXX**	20:14:14
Air Blank	IPS***	20:14:28

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 01-13-2012 TIME 2016

INTOXILYZER SERIAL # 80-001095 LOCATION MAYER

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody L7486

DPS Form Exh G-2 (Rev 05-01)

UNSTABLE SIGNAL DURING CAL CHECK - SEQUENCE TERMINATED  
NEW CAL CHECK COMPLETED - INSTRUMENT FUNCTIONING PROPERLY AND  
ACCURATELY

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
01/13/2012 20:15:35

-----  
Standard Lot#: 20910100A1  
Last Changed By: J. JAEGER #3566  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:16:04
Reference	UNS*	20:16:12
Air Blank	0.000	20:16:40

\*Unstable Signal

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
01/13/2012 20:16:43

-----  
Standard Lot#: 20910100A1  
Last Changed By: J. JAEGER #3566  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:17:13
0.100 Cal Check	0.097	20:17:34
Air Blank	0.000	20:18:03

Preventative Maintenance  
Performed -- Timer Reset