

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY #7486 AGENCY YAVAPAI CO. S.O.

DATE 10-01-2013 TIME 1707

INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution  $\pm$  100 AC.  
OR  
~~Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$~~
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 094 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody #7486

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
10/01/2013 17:07:34

-----  
Standard Lot#: 32012100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	17:08:59
0.100 Cal Check	0.094	17:09:20
Air Blank	0.000	17:09:49

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY 7486 AGENCY SAVANNAH CO. S.O.  
DATE 10-01-2013 TIME 1714  
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.  
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.  
() 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
() 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

SIGNATURE L Moody 7486

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
10/01/2013

17:14:07

---

Standard Lot#: 32012100A1  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: TEST,SQAP  
DOB: 01/01/2000  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:14:57
Diagnostic Test	Pass	17:15:32
Air Blank	0.000	17:16:00
0.100 Cal Check	0.094	17:16:20
Air Blank	0.000	17:16:50
Subject Test	0.000	17:17:24
Air Blank	ABT*	17:17:34
Air Blank	RFI**	17:17:41

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
10/01/2013

17:17:48

---

Standard Lot#: 32012100A1  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST,N  
DOB: 01/01/2000  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	17:18:41
Diagnostic Test	Pass	17:19:16
Air Blank	0.000	17:19:43
0.100 Cal Check	0.094	17:20:04
Air Blank	0.000	17:20:34
Subject Test	0.084*	17:23:41
Air Blank	0.000	17:24:12
Air Blank	0.000	17:24:40
Subject Test	XXX**	17:25:13
Air Blank	IPS***	17:25:25

\*Deficient Sample  
\*\*Invalid Sample  
    Begin new deprivation period  
\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOORS #2486 AGENCY YAVAPAI CO. SO.

DATE 10-01-2013 TIME 1726

INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.  
OR  
~~Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.094 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Moors 2486

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
10/01/2013 17:26:06

-----  
Standard Lot#: 32012100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	17:26:35
0.100 Cal Check	0.094	17:26:56
Air Blank	0.000	17:27:26

Preventative Maintenance  
Performed -- Timer Reset