

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY CAVAPAE CO. SO

DATE 12/31/2013 TIME 1709

INTOXILYZER SERIAL # 80-001095 LOCATION CCSO - MAGER

67 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR

~~Pour a standard alcohol concentration solution _____ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~

67 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"

67 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.

67 4. Air blank completed.

67 5. Calibration check completed. Test results 0. 094 AC.

67 6. Air blank completed.

67 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LMoody 7486



INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/31/2013

17:09:21

Standard Lot#: 32012100A1
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	17:09:50
0.100 Cal Check	0.094	17:10:11
Air Blank	0.000	17:10:40

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPI CO. S.O.
DATE 12/31/2013 TIME 1711
INTOXILYZER SERIAL # 80-001095 LOCATION YUCO- MAGER

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 094 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE Logan Moody 7486

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/31/2013

17:11:13

Standard Lot#: 32012100A1
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:11:59
Diagnostic Test	Pass	17:12:34
Air Blank	0.000	17:13:02
0.100 Cal Check	0.094	17:13:22
Air Blank	0.000	17:13:52
Subject Test	0.000	17:14:32
Air Blank	ABT*	17:14:39
Air Blank	RFI**	17:14:47

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence



INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/31/2013

17:14:52

Standard Lot#: 32012100A1
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:15:38
Diagnostic Test	Pass	17:16:13
Air Blank	0.000	17:16:41
0.100 Cal Check	0.095	17:17:01
Air Blank	0.000	17:17:31
Subject Test	0.045*	17:20:38
Air Blank	0.000	17:21:09
Air Blank	0.000	17:21:36
Subject Test	XXX**	17:22:04
Air Blank	IPS***	17:22:12

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.

DATE 12/31/2013 TIME 1722

INTOXILYZER SERIAL # 86-001095 LOCATION YCSO - MAGER

- (67) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution -1000 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$~~
- (67) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (67) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (67) 4. Air blank completed.
- (67) 5. Calibration check completed. Test results 0. 094 AC.
- (67) 6. Air blank completed.
- (67) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LMoody 7486



INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/31/2013 17:22:32

Standard Lot#: 32012100A1
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	17:23:01
0.100 Cal Check	0.094	17:23:21
Air Blank	0.000	17:23:51

Preventative Maintenance
Performed -- Timer Reset

RECEIVED

JAN 02 2014

YAVAPAI COUNTY ATTORNEY