

**EXHIBIT G-2**  
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**  
**ARIZONA DEPARTMENT OF PUBLIC SAFETY**  
**STANDARD QUALITY ASSURANCE PROCEDURES**  
**INTOXILYZER MODEL 8000**  
**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST LOREN MOCCOPI AGENCY YAVAPAI CO S.D.  
DATE 3/21/2014 TIME 0243  
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

- (U) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (G) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (U) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (U) 4. Air blank completed.
- (U) 5. Calibration check completed. Test results 0. 096 AC.
- (U) 6. Air blank completed.
- (U) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
06/07/2014 02:43:59

-----  
Standard Lot#: 32012100A1  
Last Changed By: L. MOODY #7486

-----  
OAS: L. MOODY #7486  
YAUAPAI CSO

-----  
Operator: L. MOODY #7486  
YAUAPAI CSO

-----  
Test                    g/210L                    Time  
-----  
Air Blank                0.000                    02:44:32  
0.100 Cal Check        0.096                    02:44:53  
Air Blank                0.000                    02:45:23

Preventative Maintenance  
Performed -- Timer Reset