

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.
DATE 01/06/2015 TIME 1439
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAYER

- BT 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC. into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- BT 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- BT 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- BT 4. Air blank completed.
- BT 5. Calibration check completed. Test results 0. 097 AC.
- BT 6. Air blank completed.
- BT 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE LMoody 7486

DPS Form Exh G-2 (Rev 05-01)

STD. CHANGE: 33913100A3
EXP: 1/1/2016
CYL: 069

INTOXILYZER 8000

Location: YCSO-MAYER

Serial Number: 80-001095

Core Version: 8105.48

01/06/2015

14:39:43

Standard Lot#: 33913100A3

Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486

YAVAPAI CSO

Operator: L. MOODY #7486

YAVAPAI CSO

Test g/210L Time

Air Blank 0.000 14:40:12
0.100 Cal Check 0.097 14:40:33
Air Blank 0.000 14:41:02

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.
DATE 01/06/2015 TIME 1443
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO-MAYER

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE LMoody 7-16

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
01/06/2015 14:43:26

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:44:13
Diagnostic Test	Pass	14:44:48
Air Blank	0.000	14:45:15
0.100 Cal Check	0.097	14:45:36
Air Blank	0.000	14:46:06
Subject Test	0.000	14:46:25
Air Blank	0.000	14:46:53
Wait	ABT*	
Air Blank	RFI**	14:47:12

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
01/06/2015 14:47:35

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	14:48:23
Diagnostic Test	Pass	14:48:58
Air Blank	0.000	14:49:25
0.100 Cal Check	0.097	14:49:46
Air Blank	0.000	14:50:15
Subject Test	0.119*	14:53:23
Air Blank	0.000	14:53:55
Air Blank	0.000	14:54:22
Subject Test	XXX**	14:55:49
Air Blank	IPS***	14:55:56

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. SO.
DATE 01/06/2015 TIME 1457
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAYER

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.
OR
~~Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34 C \pm 0.2 C.~~
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 097 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody

