

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.
DATE 2/2/2015 TIME 2238
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAYER

- (64) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (64) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (64) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (62) 4. Air blank completed.
- (64) 5. Calibration check completed. Test results 0. 097 AC.
- (64) 6. Air blank completed.
- (64) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/02/2015 22:38:26

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:38:55
0.100 Cal Check	0.097	22:39:16
Air Blank	0.000	22:39:46

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.
DATE 2/2/2015 TIME 2240
INTOXILYZER SERIAL # 80-001095 LOCATION YUSD - MAYER

(B) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (B) 1. Clock time check.
- (B) 2. Date check.

OPERATIONAL TESTS

- (B) 1. Alcohol-free subject test result 0. 000 AC.
- (B) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- (B) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- (B) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES (M) NO _____

COMMENTS

SIGNATURE Moody 7486

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/02/2015

22:40:23

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:41:07
Diagnostic Test	Pass	22:41:42
Air Blank	0.000	22:42:10
0.100 Cal Check	0.098	22:42:31
Air Blank	0.000	22:43:00
Subject Test	0.000	22:43:25
Air Blank	ABT*	22:43:35
Air Blank	RFI**	22:43:42

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/02/2015

22:43:47

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:44:35
Diagnostic Test	Pass	22:45:10
Air Blank	0.000	22:45:38
0.100 Cal Check	0.097	22:45:58
Air Blank	0.000	22:46:28
Subject Test	0.042*	22:49:35
Air Blank	0.000	22:50:06
Air Blank	0.000	22:50:34
Subject Test	XXX**	22:51:04
Air Blank	IPS***	22:51:12

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOORE AGENCY YAVAPAI CO. S.O.
DATE 2/2/2015 TIME 2252
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

- (b7) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. ~~Turn on the simulator and allow temperature to reach 34°C ± 0.2°C.~~
- (b7) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (b7) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (b7) 4. Air blank completed.
- (b7) 5. Calibration check completed. Test results 0. 097 AC.
- (b7) 6. Air blank completed.
- (b7) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moore

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/02/2015 22:52:16

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:52:45
0.100 Cal Check	0.097	22:53:06
Air Blank	0.000	22:53:35

Preventative Maintenance
Performed -- Timer Reset