

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOODY #7486 AGENCY YAVAPAI CO. S.O.  
DATE 11/24/2015 TIME 1845  
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAYER

- (1) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.  
OR  
~~Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$~~
- (2) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (3) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (4) 4. Air blank completed.
- (5) 5. Calibration check completed. Test results 0. 097 AC.
- (6) 6. Air blank completed.
- (7) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody #7486

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
11/24/2015 18:45:03

-----  
Standard Lot#: 33913100A3  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:45:32
0.100 Cal Check	0.097	18:45:53
Air Blank	0.000	18:46:22

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOORE 4748 AGENCY YAVAPAI CO. S.O  
DATE 11/24/2015 TIME 1846  
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO-MAYER

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES  NO

COMMENTS

750 psi

SIGNATURE

Logan Moore 4748

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
11/24/2015 18:46:33

---

Standard Lot#: 33913100A3  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:47:19
Diagnostic Test	Pass	18:47:54
Air Blank	0.000	18:48:21
0.100 Cal Check	0.098	18:48:42
Air Blank	0.000	18:49:11
Subject Test	0.000	18:49:41
Air Blank	ABT*	18:49:50
Air Blank	RFI**	18:49:57

\*Sequence Aborted  
\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
11/24/2015 18:50:16

---

Standard Lot#: 33913100A3  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/2000  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:51:05
Diagnostic Test	Pass	18:51:40
Air Blank	0.000	18:52:07
0.100 Cal Check	0.097	18:52:28
Air Blank	0.000	18:52:57
Subject Test	0.045*	18:56:05
Air Blank	0.000	18:56:36
Air Blank	0.000	18:57:03
Subject Test	XXX**	18:57:34
Air Blank	IPS***	18:57:42

\*Deficient Sample  
\*\*Invalid Sample  
    Begin new deprivation period  
\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

**EXHIBIT G-2**  
**THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)**  
**ARIZONA DEPARTMENT OF PUBLIC SAFETY**  
**STANDARD QUALITY ASSURANCE PROCEDURES**  
**INTOXILYZER MODEL 8000**  
**STANDARD CALIBRATION CHECK PROCEDURE**

QA SPECIALIST LOGAN MOODY AGENCY YAVAPAI CO. S.O.  
DATE 11/24/2015 TIME 1859  
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO-MAGER

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
~~Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 097 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody 7-486

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
11/24/2015 18:59:14

-----  
Standard Lot#: 33913100A3  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:59:43
0.100 Cal Check	0.097	19:00:04
Air Blank	0.000	19:00:33

Preventative Maintenance  
Performed -- Timer Reset