

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOORE AGENCY YAVAPAI CO. S.O.
DATE 12/21/15 TIME 2209
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - MAGER

- (M) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a ~~clean dry simulator~~ and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (M) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(M) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(M) 4. Air blank completed.
(M) 5. Calibration check completed. Test results 0. 097 AC.
(M) 6. Air blank completed.
(M) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moore

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/21/2015 22:09:44

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:10:13 |
| 0.100 Cal Check | 0.097 | 22:10:34 |
| Air Blank | 0.000 | 22:11:03 |

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST LOGAN MOORE AGENCY YAVAPAI CO SO.
DATE 12/21/2015 TIME 2212
INTOXILYZER SERIAL # 80-001095 LOCATION YUCO - MAYER

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

700 psi

SIGNATURE Logan Moore

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/21/2015

22:12:05

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:12:53 |
| Diagnostic Test | Pass | 22:13:28 |
| Air Blank | 0.000 | 22:13:56 |
| 0.100 Cal Check | 0.097 | 22:14:16 |
| Air Blank | 0.000 | 22:14:46 |
| Subject Test | 0.000 | 22:15:47 |
| Air Blank | ABT* | 22:15:55 |
| Air Blank | RFI** | 22:16:03 |

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/21/2015 22:16:30

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/2000
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:17:15 |
| Diagnostic Test | Pass | 22:17:50 |
| Air Blank | 0.000 | 22:18:18 |
| 0.100 Cal Check | 0.097 | 22:18:39 |
| Air Blank | 0.000 | 22:19:08 |
| Subject Test | 0.058* | 22:22:16 |
| Air Blank | 0.000 | 22:22:47 |
| Air Blank | 0.000 | 22:23:14 |
| Subject Test | XXX** | 22:23:58 |
| Air Blank | IPS*** | 22:24:06 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST LOGAN MOORE AGENCY YAVAPAI CO. SO.
DATE 12/21/2015 TIME 2225
INTOXILYZER SERIAL # 80-001095 LOCATION YUCO - MAGER

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution: 100 AC.
OR
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 0917 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moore

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
12/21/2015 22:25:50

Standard Lot#: 33913100A3
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 22:26:18 |
| 0.100 Cal Check | 0.097 | 22:26:39 |
| Air Blank | 0.000 | 22:27:08 |

Preventative Maintenance
Performed -- Timer Reset