

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SEAN CHUPP AGENCY YCSO

DATE 9/22/14 TIME 1821

INTOXILYZER SERIAL # 80-001095 LOCATION MAYER OFFICE

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  E39/9808

INTOXILYZER 8000

Location: YCSO-MAYER

Serial Number: 80-001095

Core Version: 8105.48

05/22/2016

18:21:36

Standard Lot#: 15615100A1

Last Changed By: L. MOODY #7486

QAS: S. CHUPP #9808

YAVAPAI CSO

Operator: S. CHUPP #9808

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	18:22:10
0.100 Cal Check	0.099	18:22:31
Air Blank	0.000	18:23:00

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST SEAN CHUPP AGENCY YCSO
DATE 9/22/16 TIME 1841
INTOXILYZER SERIAL # 80-001095 LOCATION MAYER OFFICE

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a ~~standard alcohol concentration solution~~ AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  E39/9808

INTOXILYZER 8000
Location: YCSO-MAYER.
Serial Number: 80-001095
Core Version: 8105.48
05/22/2016 18:41:22

Standard Lot#: 15615100A1
Last Changed By: L. MOODY #7486

QAS: S. CHUPP #9808
YAVAPAI CSO

Operator: S. CHUPP #9808
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	18:41:52
0.100 Cal Check	0.098	18:42:13
Air Blank	0.000	18:42:42

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST SEAN CHUPP AGENCY YCSO

DATE 3/22/16 TIME 1824

INTOXILYZER SERIAL # 80-001095 LOCATION MAYER OFFICE

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS


- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS
950 psi

SIGNATURE  E39/9808

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
05/22/2016

18:26:33

Standard Lot#: 15615100A1
Last Changed By: L. MOODY #7486

QAS: S. CHUPP #9808
YAVAPAI CSO

Operator: S. CHUPP #9808
YAVAPAICSO

Subject: SQAP,TEST
DOB: 01/01/1991
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:27:28
Diagnostic Test	Pass	18:28:03
Air Blank	0.000	18:28:30
0.100 Cal Check	0.099	18:28:51
Air Blank	0.000	18:29:21
Subject Test	0.000	18:29:40
Air Blank	RFI*	18:29:51
Air Blank	IPS**	18:30:08

*RFI Detect

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO-MAYER

Serial Number: 80-001095

Core Version: 8105.48

05/22/2016

18:30:45

Standard Lot#: 15615100A1

Last Changed By: L. MOODY #7486

QAS: S. CHUPP #9808

YAVAPAI CSO

Operator: S. CHUPP #9808

YAVAPAI CSO

Subject: SQAP,TEST

DOB: 01/01/1991

Sex: M

Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	ABT*	18:31:33
Air Blank	0.000	18:32:02

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
05/22/2016

18:32:07

Standard Lot#: 15615100A1
Last Changed By: L. MOODY #7486

QAS: S. CHUPP #9808
YAVAPAI CSO

Operator: S. CHUPP #9808
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/1991
Sex: M Weight: 180

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	18:33:01
Diagnostic Test	Pass	18:33:36
Air Blank	0.000	18:34:03
0.100 Cal Check	0.098	18:34:24
Air Blank	0.000	18:34:53
Subject Test	0.102*	18:38:01
Air Blank	0.000	18:38:32
Air Blank	0.000	18:39:00
Subject Test	XXX**	18:39:19
Air Blank	0.000	18:39:55

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence