

16-045869

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
12/28/2016 16:35:59

-----  
Standard Lot#: 15615100A1  
Last Changed By: L. MOODY #7486

-----  
QAS: L. MOODY #7486  
YAVAPAI CSO

-----  
Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	16:36:28
0.100 Cal Check	0.098	16:36:49
Air Blank	0.000	16:37:18

Preventative Maintenance  
Performed -- Timer Reset

16-045869

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
12/28/2016

16:51:20

---

Standard Lot#: 15615100A1  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:52:12
Diagnostic Test	Pass	16:52:47
Air Blank	0.000	16:53:14
0.100 Cal Check	0.099	16:53:35
Air Blank	0.000	16:54:05
Subject Test	0.161*	16:57:12
Air Blank	0.000	16:57:44
Air Blank	0.000	16:58:11
Subject Test	XXX**	16:58:36
Air Blank	IPS***	16:58:44

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

16-045869

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
12/28/2016 16:58:58

-----  
Standard Lot#: 15615100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	16:59:26
0.100 Cal Check	0.098	16:59:47
Air Blank	0.000	17:00:17

Preventative Maintenance  
Performed -- Timer Reset

16-045869

INTOXILYZER 8000  
Location: YCSO-MAYER  
Serial Number: 80-001095  
Core Version: 8105.48  
12/28/2016 16:47:53

---

Standard Lot#: 15615100A1  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	16:48:46
Diagnostic Test	Pass	16:49:21
Air Blank	0.000	16:49:48
0.100 Cal Check	0.099	16:50:09
Air Blank	0.000	16:50:39
Subject Test	0.000	16:50:58
Air Blank	ABT*	16:51:07
Air Blank	RFI**	16:51:14

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

16-045852

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
12/28/2016 18:36:20

---

Standard Lot#: 23016100A1  
Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
YAVAPAI CSO

---

Operator: L. MOODY #7486  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 01/01/1990  
Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:37:05
Diagnostic Test	Pass	18:37:40
Air Blank	0.000	18:38:07
0.100 Cal Check	0.100	18:38:28
Air Blank	0.000	18:38:57
Subject Test	0.101*	18:42:05
Air Blank	0.000	18:42:36
Air Blank	0.000	18:43:04
Subject Test	XXX**	18:43:34
Air Blank	IPS***	18:43:41

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

16-045882

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
12/28/2016 18:30:05

-----  
Standard Lot#: 23016100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:30:34
0.100 Cal Check	0.101	18:30:55
Air Blank	0.000	18:31:24

Preventative Maintenance  
Performed -- Timer Reset

16-045882

INTOXILYZER 8000  
Location: YCSO - BLACK CANYON  
Serial Number: 80-000289  
Core Version: 8105.48  
12/28/2016 18:43:57

-----  
Standard Lot#: 23016100A1  
Last Changed By: L. MOODY #7486  
-----

QAS: L. MOODY #7486  
YAVAPAI CSO  
-----

Operator: L. MOODY #7486  
YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	18:44:25
0.100 Cal Check	0.100	18:44:46
Air Blank	0.000	18:45:15

Preventative Maintenance  
Performed -- Timer Reset

INTOXILYZER 8000  
 Location: YCSO - BLACK CANYON  
 Serial Number: 80-000289  
 Core Version: 8105.48  
 12/28/2016 18:32:58

---

Standard Lot#: 23016100A1  
 Last Changed By: L. MOODY #7486

---

QAS: L. MOODY #7486  
 YAVAPAI CSO

---

Operator: L. MOODY #7486  
 YAVAPAI CSO

---

Subject: SQAP, TEST  
 DOB: 01/01/1990  
 Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	18:33:43
Diagnostic Test	Pass	18:34:17
Air Blank	0.000	18:34:45
0.100 Cal Check	0.101	18:35:06
Air Blank	0.000	18:35:35
Subject Test	0.000	18:35:52
Air Blank	ABT*	18:36:03
Air Blank	RFI**	18:36:11

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**