

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
DATE 02/01/2018 TIME 1325
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

- (10) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution AC, into a ~~clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- (10) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (10) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (10) 4. Air blank completed.
- (10) 5. Calibration check completed. Test results 0. 099 AC.
- (10) 6. Air blank completed.
- (10) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody 7486

18-003927

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/01/2018 13:25:07

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 13:25:35 |
| 0.100 Cal Check | 0.099 | 13:25:56 |
| Air Blank | 0.000 | 13:26:26 |

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
DATE 02/01/2018 TIME 1337
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

- (LM) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 100 AC.
OR
~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- (LM) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (LM) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (LM) 4. Air blank completed.
- (LM) 5. Calibration check completed. Test results 0. 098 AC.
- (LM) 6. Air blank completed.
- (LM) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody #7486

18-003927

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/01/2018 13:37:46

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 13:38:14 |
| 0.100 Cal Check | 0.098 | 13:38:35 |
| Air Blank | 0.000 | 13:39:05 |

Preventative Maintenance
Performed -- Timer Reset

18-003927

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office

DATE 02/01/2018 TIME 1326

INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS 950 psi

SIGNATURE Logan Moody 7486

18-003927

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/01/2018

13:26:45

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 13:27:43 |
| Diagnostic Test | Pass | 13:28:18 |
| Air Blank | 0.000 | 13:28:45 |
| 0.100 Cal Check | 0.098 | 13:29:06 |
| Air Blank | 0.000 | 13:29:35 |
| Subject Test | 0.000 | 13:30:02 |
| Air Blank | ABT* | 13:30:13 |
| Air Blank | RFI** | 13:30:20 |

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

18-003927

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
02/01/2018

13:30:26

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

| Test | g/210L | Time |
|-----------------|--------|----------|
| Air Blank | 0.000 | 13:31:13 |
| Diagnostic Test | Pass | 13:31:48 |
| Air Blank | 0.000 | 13:32:15 |
| 0.100 Cal Check | 0.098 | 13:32:36 |
| Air Blank | 0.000 | 13:33:06 |
| Subject Test | 0.105* | 13:36:13 |
| Air Blank | 0.000 | 13:36:45 |
| Air Blank | 0.000 | 13:37:12 |
| Subject Test | XXX** | 13:37:27 |
| Air Blank | IPS*** | 13:37:34 |

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence