

18-032379

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
DATE 09/01/2018 TIME 0813
INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

- (u) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (u) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (u) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (u) 4. Air blank completed.
- (u) 5. Calibration check completed. Test results 0. 099 AC.
- (u) 6. Air blank completed.
- (u) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody #7486

18-032379

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
09/01/2018 08:13:58

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	08:14:27
0.100 Cal Check	0.099	08:14:48
Air Blank	0.000	08:15:17

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office

DATE 09/01/2018 TIME 0815

INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- 1. Clock time check.
- 2. Date check.

OPERATIONAL TESTS

- 1. Alcohol-free subject test result 0. 000 AC.
- 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS 850 psi

SIGNATURE Logan Moody #7486

18-032379

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
09/01/2018 08:15:30

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 01/01/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:16:23
Diagnostic Test	Pass	08:16:58
Air Blank	0.000	08:17:25
0.100 Cal Check	0.098	08:17:46
Air Blank	0.000	08:18:16
Subject Test	0.000	08:18:44
Air Blank	ABT*	08:18:53
Air Blank	RFI**	08:19:01

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
 Location: YCSO-MAYER
 Serial Number: 80-001095
 Core Version: 8105.48
 09/01/2018 08:19:20

Standard Lot#: 06717100A5
 Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
 YAVAPAI CSO

Operator: L. MOODY #7486
 YAVAPAI CSO

Subject: SQAP, TEST
 DOB: 01/01/1990
 Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:20:11
Diagnostic Test	Pass	08:20:46
Air Blank	0.000	08:21:13
0.100 Cal Check	0.098	08:21:34
Air Blank	0.000	08:22:03
Subject Test	0.100*	08:25:11
Air Blank	0.000	08:25:42
Air Blank	0.000	08:26:09
Subject Test	XXX**	08:26:29
Air Blank	IPS***	08:26:36

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

18-032379

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)
ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Logan Moody #7486 AGENCY Yavapai County Sheriff's Office
 DATE 09/01/2018 TIME 0828
 INTOXILYZER SERIAL # 80-001095 LOCATION YCSO - Mayer

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution. 100 AC.
 OR
~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 098 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Logan Moody #7486

18-032379

INTOXILYZER 8000
Location: YCSO-MAYER
Serial Number: 80-001095
Core Version: 8105.48
09/01/2018 08:28:59

Standard Lot#: 06717100A5
Last Changed By: L. MOODY #7486

QAS: L. MOODY #7486
YAVAPAI CSO

Operator: L. MOODY #7486
YAVAPAI CSO

Test g/210L Time

Air Blank 0.000 08:29:27
0.100 Cal Check 0.098 08:29:48
Air Blank 0.000 08:30:17

Preventative Maintenance
Performed -- Timer Reset