

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Grover AGENCY C.V.M.O.
DATE 11-29-08 TIME 11:56
INTOXILYZER SERIAL # FO-001096 OPS LOCATION C.V.M.O.

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 097 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Russell Grover #22

DPS Form Exh G-2 (Rev 05-01)

Lot # 706502 I
EXP DATE 3-9-09

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Russell Grover AGENCY C.V.M.O.

DATE 11-29-08 TIME 11:56

INTOXILYZER SERIAL # 80-081096 DPS LOCATION C.V.M.O.

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS Nothing wrong, appears to be working properly.

SIGNATURE Russell Grover #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
11/29/2008 11:54:59

Standard Lot#: 706502I
Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	11:55:26
Cal Check	0.097	11:55:45
Air Blank	0.000	11:56:15



11-29-08

Operator's Signature

Lot # 706502I
EXP DATA 3-9-09

INTOXILYZER 8000

Location: CVMO

Serial Number : 80-001096

11/29/2008

11:48:32

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Operator: R. GROVER #22

CAMP VERDE MO

Subject: SQAP,TEST

DOB: 08/19/1971

Sex: M

Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:49:31
Diagnostic Test	Pass	11:49:58
Air Blank	0.000	11:50:26
0.100 Cal Check	0.097	11:50:45
Air Blank	0.000	11:51:14
Subject Test	0.000	11:51:34
Air Blank	0.000	11:52:01
Five Minute Wait	ABT*	
Air Blank	0.000	11:52:39

*Sequence Aborted

Not a Successfully Completed Test Sequence

R. Grover #22

11/29/08

INTOXILYZER 8000

Location: CVMG

Serial Number : 80-001096

11/29/2008

12:00:20

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Operator: R. GROVER #22

CAMP VERDE MO

Subject: SOAP,TEST

DOB: 08/19/1971

Sex: M

Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	12:01:43
Air Blank	0.000	12:02:11

*Improper Sample

Not a Successfully Completed Test Sequence

Ronald M #22

11/29/08

INTOXILYZER 8000
Location: CVMO
Serial Number : 80-001096
11/29/2008

12:03:40

Standard Lot#: 706502I
Last Changed By: S. MARTINKO #25
QAS: R. GROVER #22
CAMP VERDE MO
Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225
15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:04:19
Air Blank	0.000	12:04:47

*RFI Detect

Not a Successfully Completed Test Sequence

Ronald M #22

11/29/08

INTOXILYZER 8000

Location: CVMO

Serial Number : 80-001096

11/29/2008

12:19:08

Standard Lot#: 706502I

Last Changed By: S. MARTINKO

#25

OAS: R. GROVER

#22

CAMP VERDE MO

Operator: R. GROVER

#22

CAMP VERDE MO

Subject: SQAP,TEST

DOB: 08/19/1971

Sex: M

Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:20:01
Diagnostic Test	Pass	12:20:25
Air Blank	0.000	12:20:56
0.100 Cal Check	0.097	12:21:15
Air Blank	0.000	12:21:44
Subject Test	0.284*	12:24:53
Air Blank	0.000	12:25:25
Air Blank	0.000	12:25:53
Subject Test	XXX**	12:26:36
Air Blank	0.000	12:27:11

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

R. Grover #22

11/29/08

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001096

11/29/2008

12:30:41

Standard Lot#: 7065021

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Test g/210L Time

Air Blank 0.000 12:31:09
Cal Check 0.097 12:31:28
Air Blank 0.000 12:31:57

R. Grover #22 11/29/08

Operator's Signature

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Waver AGENCY C.V.M.O.
DATE 11-29-08 TIME 1243
INTOXILYZER SERIAL # 80-001736 DPS LOCATION C.V.M.O.

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 11/29/08
#22

DPS Form Exh G-2 (Rev 05-01)

Lot # 706502 I
EXP DATE = 3-9-09

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Russell Grover AGENCY C.V.M.O.
DATE 11-29-08 TIME 1315
INTOXILYZER SERIAL # 80-001736 DPS LOCATION C.V.M.O.

(✓) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (✓) 1. Clock time check.
(✓) 2. Date check.

OPERATIONAL TESTS

- (✓) 1. Alcohol-free subject test result 0. 000 AC.
(✓) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(✓) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(✓) 4. Standard Calibration Check standard 0. 100 AC. Result 0. 096 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Appears to be working properly

SIGNATURE Russell Grover #22 11-29-08

INTOXILYZER 8000

Location: DPS

Serial Number: 80-001736

11/29/2008

12:42:43

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	12:43:11
Cal Check	0.097	12:43:30
Air Blank	0.000	12:43:59

R. Grover #22 11/29/08

Operator's Signature

Lot # 706502 I

EXP DATE = 3-9-09

INTOXILYZER 8000

Location: DPS

Serial Number : 80-001736

11/29/2008

12:46:00

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Operator: R. GROVER #22

CVMO

Subject: SQAP,TEST

DOB: 08/19/1971

Sex: M

Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	12:47:01
Diagnostic Test	Pass	12:47:28
Air Blank	0.000	12:47:55
0.100 Cal Check	0.097	12:48:14
Air Blank	0.000	12:48:44
Subject Test	0.000	12:49:09
Air Blank	0.000	12:49:36
Five Minute Wait	ABT*	
Air Blank	0.000	12:50:23

*Sequence Aborted

Not a Successfully Completed Test Sequence

R. Grover #22 11/29/08

INTOXILYZER 8000

Location: DPS

Serial Number : 80-001736

11/29/2008

12:52:03

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Operator: R. GROVER #22

CAMP VERDE MO

Subject: SQAP, TEST

DOB: 08/19/1971

Sex: M

Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	12:52:35
Air Blank	0.000	12:53:04

*Improper Sample

Not a Successfully Completed Test Sequence

Ronald M #22

11/29/08

INTOXILYZER 8000
Location: DPS
Serial Number : 80-001736
11/29/2008 12:53:52

Standard Lot#: 706502I
Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	12:54:20
Air Blank	0.000	12:54:48

*RFI Detect

Not a Successfully Completed Test Sequence

Ronald H. Herz 11/29/08

INTOXILYZER 8000
Location: DPS
Serial Number : 80-001736
11/29/2008 13:02:55

Standard Lot#: 706502I
Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	13:03:40
Diagnostic Test	Pass	13:04:08
Air Blank	0.000	13:04:35
0.100 Cal Check	0.097	13:04:54
Air Blank	0.000	13:05:23
Subject Test	0.283*	13:08:32
Air Blank	0.000	13:09:04
Air Blank	0.000	13:09:32
Subject Test	XXX**	13:09:54
Air Blank	0.000	13:10:35

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

R. Grover #22 11/29/08

INTOXILYZER 8000

Location: DPS

Serial Number: 80-001736

11/29/2008

13:14:18

Standard Lot#: 706502I

Last Changed By: S. MARTINKO #25

QAS: R. GROVER #22

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	13:14:46
Cal Check	0.096	13:15:05
Air Blank	0.000	13:15:34

 11/29/08

Operator's Signature

Lot # 706502 I

SKY DATE 3-9-09