

31 DAY

COPY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Crowe AGENCY CVMO
DATE 4-28-10 TIME 0615
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. M. #22

DPS Form Exh G-2 (Rev 05-01)

EXP DATE 12-4-2010

LOT # ~~12-4-2010~~
8329016

Camp Verde Marshal's Office
This report and ALL attachments
included with this file are for
LAW ENFORCEMENT USE ONLY
DISSEMINATION IS RESTRICTED

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
04/28/2010 06:14:58

Standard Lot#: 832901G
Last Changed By: R. GROVER #22

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	06:15:38
0.100 Cal Check	0.098	06:15:58
Air Blank	0.000	06:16:30

Preventative Maintenance
Performed -- Timer Reset

R. M. #22

4/28/10