

31 DAY

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Grover AGENCY CVMO  
DATE 7-25-10 TIME 1631  
INTOXILYZER SERIAL # 80-001096 DPS LOCATION CVMO

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Grover #22

DPS Form Exh G-2 (Rev 05-01)

ETP DATE 12-4-10

LOT # 8329016

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001096

Core Version: 8105.48

07/25/2010

16:30:50

-----  
Standard Lot#: 832901G

Last Changed By: R. GROVER #22  
-----

QAS: R. GROVER #22

CAMP VERDE MO  
-----

Operator: R. GROVER #22

CAMP VERDE MO  
-----

Test	g/210L	Time
Air Blank	0.000	16:31:25
0.100 Cal Check	0.098	16:31:46
Air Blank	0.000	16:32:18

Preventative Maintenance  
Performed -- Timer Reset

*R. Grover #22*

7-25-10

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Groves AGENCY CVMO  
DATE 7-25-10 TIME 1711 "  
INTOXILYZER SERIAL # 80-001736 DPS LOCATION CVMO

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 1100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 097 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Groves #22

DPS Form Exh G-2 (Rev 05-01)

~~ETA~~ DATE 12-4-10  
LOT # 832901 F

INTOXILYZER 8000

Location: DPS

Serial Number: 80-001736

Core Version: 8105.48

07/25/2010

17:10:59

-----  
Standard Lot#: 832901G

Last Changed By: R. GROVER #22  
-----

QAS: R. GROVER #22

CAMP VERDE MO  
-----

Operator: R. GROVER #22

CAMP VERDE MO  
-----

Test	g/210L	Time
Air Blank	0.000	17:11:35
0.100 Cal Check	0.097	17:11:56
Air Blank	0.000	17:12:25

Preventative Maintenance  
Performed -- Timer Reset

*R. Grover #22*

*LOT # 832901 G*

*EXP DATE 12-4-10*