

90 DAY

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000.
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RUSSELL GROVER AGENCY CVMU
DATE 2-23-11 TIME 0409
INTOXILYZER SERIAL # 80-001096 DPS LOCATION CVMU

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0 .099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Grover #22

DPS Form Exh G-2 (Rev 05-01)

LOT # 767153
EXP DATE 5-21-12

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RUSSETI WAVE AGENCY CVMO
DATE 2-23-11 TIME 0433
INTOXILYZER SERIAL # 80-001096DPS LOCATION CVMO

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS WORKS PROPERLY

SIGNATURE R. H. #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011

04:08:45

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	04:09:23
0.100 Cal Check	0.099	04:09:43
Air Blank	0.000	04:10:13

Preventative Maintenance
Performed -- Timer Reset

R. Grover #22

2/23/11

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011 04:13:26

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:14:19
Diagnostic Test	Pass	04:14:54
Air Blank	0.000	04:15:21
0.100 Cal Check	0.099	04:15:42
Air Blank	0.000	04:16:11
Subject Test	0.000	04:16:33
Air Blank	0.000	04:17:01
Wait	ABT*	
Air Blank	0.000	04:17:54

*Sequence Aborted

Not a Successfully Completed Test Sequence

R. Grover #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011 04:19:01

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	04:19:34
Air Blank	0.000	04:20:02

*Improper Sample

Not a Successfully Completed Test Sequence

R. Grover #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011

04:20:39

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	04:21:20
Air Blank	0.000	04:21:48

*RFI Detect

Not a Successfully Completed Test Sequence

R. M. #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011 04:23:21

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 225

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	04:24:20
Diagnostic Test	Pass	04:24:55
Air Blank	0.000	04:25:22
0.100 Cal Check	0.099	04:25:43
Air Blank	0.000	04:26:12
Subject Test	0.103*	04:29:20
Air Blank	0.000	04:29:51
Air Blank	0.000	04:30:19
Subject Test	XXX**	04:31:06
Air Blank	0.000	04:31:39

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

R. Grover #22

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
02/23/2011 04:32:17

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	04:32:46
0.100 Cal Check	0.098	04:33:07
Air Blank	0.000	04:33:38

Preventative Maintenance
Performed -- Timer Reset

R. Grover #22