

90 DAY TEST

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Russell Crowe AGENCY CVMC
DATE 5-19-11 TIME 1556
INTOXILYZER SERIAL # 80-1096 LOCATION CVMC

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Crowe

DPS Form Exh G-2 (Rev 05-01)

Lot # 767153
EX DATE 5-21-12

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Russell Crover AGENCY CVMO
DATE 5-19-11 TIME 1617
INTOXILYZER SERIAL # 80-1096 LOCATION CVMO

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
- () 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
- () 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- () 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- () 4. Standard Calibration Check standard 0. 100 AC. Result 0. 097 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

WORKS PROPERLY

SIGNATURE R. Crover

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/19/2011 15:55:05

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	15:55:43
0.100 Cal Check	0.098	15:56:04
Air Blank	0.000	15:56:35

Preventative Maintenance
Performed -- Timer Reset

R. Grover 5/19/11

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/19/2011 15:57:45

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	15:59:05
Diagnostic Test	Pass	15:59:41
Air Blank	0.000	16:00:08
0.100 Cal Check	0.098	16:00:29
Air Blank	0.000	16:01:00
Subject Test	0.000	16:01:14
Air Blank	0.000	16:01:41
Wait	ABT*	
Air Blank	0.000	16:02:22

*Sequence Aborted

Not a Successfully Completed Test Sequence

R. Grover 5/19/11

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/19/2011 16:04:01

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP, TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	IPS*	16:04:49
Air Blank	0.000	16:05:17

*Improper Sample

R. Grover

5/19/11

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/19/2011 16:05:49

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	16:06:22
Air Blank	0.000	16:06:51

*RFI Detect

R. Grover 5/19/11

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/19/2011

16:07:27

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:08:15
Diagnostic Test	Pass	16:08:50
Air Blank	0.000	16:09:18
0.100 Cal Check	0.098	16:09:39
Air Blank	0.000	16:10:08
Subject Test	0.261*	16:13:16
Air Blank	0.000	16:13:51
Air Blank	0.000	16:14:18
Subject Test	XXX**	16:14:37
Air Blank	0.000	16:15:10

*Deficient Sample

**Invalid Sample

Begin new deprivation period

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: CVMO

Serial Number: 80-001096

Core Version: 8105.48

05/19/2011

16:16:54

Standard Lot#: 767153

Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22

CAMP VERDE MO

Operator: R. GROVER #22

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	16:17:24
0.100 Cal Check	0.097	16:17:44
Air Blank	0.000	16:18:14

Preventative Maintenance
Performed -- Timer Reset

R.R. 5/19/11