

31 DAY TEST

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RUSSELL GROVER AGENCY CVMO
DATE 4-24-12 TIME 1640
INTOXILYZER SERIAL # 80-007736DPS LOCATION CVMO

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0. 098 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Grover 0-12

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
04/24/2012 16:39:03

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	16:39:46
0.100 Cal Check	0.098	16:40:07
Air Blank	0.000	16:40:38

Preventative Maintenance
Performed -- Timer Reset

R. Grover 4-24-12