

90 DAY TEST BOTH

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RUSSELL GROVED AGENCY CVMO
DATE 5-11-12 TIME 0800
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. 098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE R. Groved

Lot # 767153

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Russell Givver AGENCY CUMM
DATE 5-11-12 TIME 0830
INTOXILYZER SERIAL # 80-001096 LOCATION CUMM

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 000 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Works Properly

SIGNATURE

R. Givver 5-11-12

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/11/2012 07:59:12

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	07:59:43
0.100 Cal Check	0.098	08:00:04
Air Blank	0.000	08:00:36

Preventative Maintenance
Performed -- Timer Reset

R.N

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/11/2012

08:01:53

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:03:00
Diagnostic Test	Pass	08:03:34
Air Blank	0.000	08:04:02
0.100 Cal Check	0.098	08:04:23
Air Blank	0.000	08:04:52
Subject Test	0.000	08:05:20
Air Blank	0.000	08:05:47
Wait	ABT*	
Air Blank	0.000	08:07:33

*Sequence Aborted



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/11/2012 08:14:29

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	RFI*	08:15:13
Air Blank	0.000	08:15:41

*RFI Detect

R. Grover 10-12

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/11/2012

08:18:11

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Subject: SQAP,TEST
DOB: 08/19/1971
Sex: M Weight: 250


15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	08:19:25
Diagnostic Test	Pass	08:20:00
Air Blank	0.000	08:20:27
0.100 Cal Check	0.098	08:20:48
Air Blank	0.000	08:21:17
Subject Test	0.096*	08:24:25
Air Blank	0.000	08:24:56
Air Blank	0.000	08:25:24
Subject Test	XXX**	08:25:45
Air Blank	0.000	08:26:18

*Deficient Sample

**Invalid Sample

Begin new deprivation period



Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CVMO
Serial Number: 80-001096
Core Version: 8105.48
05/11/2012 08:29:28

Standard Lot#: 767153
Last Changed By: J. SLOWINSKI #6101

QAS: R. GROVER #22
CAMP VERDE MO

Operator: R. GROVER #22
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	08:29:58
0.100 Cal Check	0.098	08:30:19
Air Blank	0.000	08:30:50

Preventative Maintenance
Performed -- Timer Reset

