

90 DAY TEST BOTH

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)  
ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST RUSSELL GROVED AGENCY CVMO  
DATE 8-16-12 TIME \_\_\_\_\_  
INTOXILYZER SERIAL # 80-001096 DPS LOCATION CVMO

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution \_\_\_\_\_ AC OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a simulator. Ensure that a tight seal has been made. Turn on the simulator. 0.2°C
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "V" made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0. \_\_\_\_\_ AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

90 DAY TEST  
NOT DONE DUE TO  
GAS CYLINDER WAS LOW  
ORDERED NEW ONE  
ON 8-21-12

SIGNATURE R. W D-12

DPS Form Exh G-2 (Rev 05-01)

\* 90 DAY TEST WAS DONE ON 11-7-12 WHEN MACHINE WAS PICKED UP FROM DPS ALCOHOL UNIT  
R. W D-12

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST RUSSELL GROVER AGENCY CVMO

DATE 8-16-12 TIME \_\_\_\_\_

INTOXILYZER SERIAL # 80-001096 DPS LOCATION CVMO

(  ) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (  ) 1. Clock time check.  
(  ) 2. Date check.

OPERATIONAL TESTS

- (  ) 1. Alcohol-free subject test result 0. 000 AC.  
(  ) 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
(  ) 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
(  ) 4. Standard Calibration Check standard 0. 100 AC. Result 0. \_\_\_\_\_ AC.

Instrument is operating properly and accurately. YES \_\_\_\_\_ NO \_\_\_\_\_

COMMENTS

HAD TO WAIT FOR NEW GAS CYLINDER TO ARRIVE  
ORDERED ON 8-21-12

SIGNATURE R. Grover