

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY CVMO  
DATE 10/21/15 TIME 1322  
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Evidence Room

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(X) 4. Air blank completed.  
(X) 5. Calibration check completed. Test results 0. 098 AC.  
(X) 6. Air blank completed.  
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] CV23

INTOXILYZER 8000  
Location: CAMP VERDE MO  
Serial Number: 80-001096  
Core Version: 8105.48  
10/21/2015 13:22:20

-----  
Standard Lot#: 33814100A1  
Last Changed By: D. HUGHES #53  
-----

QAS: D. JACOBS #CV23  
CAMP VERDE MO  
-----

Operator: D. JACOBS #CV23  
CAMP VERDE MO  
-----

Test	g/210L	Time
Air Blank	0.000	13:22:51
0.100 Cal Check	0.098	13:23:12
Air Blank	0.000	13:23:41

Preventative Maintenance  
Performed -- Timer Reset

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST D. Jacobs AGENCY CVMO  
DATE 10/21/15 TIME 1307  
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Evidence Room

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.  
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC:  
 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
 4. Standard Calibration Check standard 0. 100 AC. Result 0. 098 AC.

Instrument is operating properly and accurately. YES  NO


COMMENTS

---

---

---

---

SIGNATURE  CV23

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

10/21/2015

13:07:23

---

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

---

QAS: D. JACOBS #CV23

CAMP VERDE MO

---

Operator: D. JACOBS #CV23

CAMP VERDE MO

---

Subject: TEST,TEST

DOB: 10/21/1990

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:08:40
Diagnostic Test	Pass	13:09:14
Air Blank	0.000	13:09:41
0.100 Cal Check	0.099	13:10:02
Air Blank	0.000	13:10:31
Subject Test	0.000	13:10:56
Air Blank	0.000	13:11:23
Wait	ABT*	
Air Blank	IPS**	13:11:37

\*Sequence Aborted

\*\*Improper Sample

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

10/21/2015

13:12:26

---

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

---

OAS: D. JACOBS #CV23

CAMP VERDE MO

---

Operator: D. JACOBS #CV23

CAMP VERDE MO

---

Subject: TEST,TEST

DOB: 10/21/1990

Sex: M Weight: 200

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	13:13:26
Diagnostic Test	Pass	13:14:00
Air Blank	0.000	13:14:27
0.100 Cal Check	0.098	13:14:48
Air Blank	0.000	13:15:18
Subject Test	0.095*	13:18:25
Air Blank	0.000	13:18:57
Air Blank	0.000	13:19:24
Subject Test	XXX**	13:19:49
Air Blank	RFI***	13:20:02

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL, 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY CVMO  
DATE 10/21/15 TIME 1304  
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Evidence Room

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"  
(X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.  
(X) 4. Air blank completed.  
(X) 5. Calibration check completed. Test results 0. 099 AC.  
(X) 6. Air blank completed.  
(X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV23

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

10/21/2015

13:04:49

-----  
Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53  
-----

QAS: D. JACOBS #CV23

CAMP VERDE MO  
-----

Operator: D. JACOBS #CV23

CAMP VERDE MO  
-----

Test	g/210L	Time
Air Blank	0.000	13:05:33
0.100 Cal Check	0.099	13:05:54
Air Blank	0.000	13:06:23

Preventative Maintenance  
Performed -- Timer Reset