

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's ofc
DATE 2-10-16 TIME 1601
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO-DWI-Trailer

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (X) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (X) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (X) 4. Air blank completed.
- (X) 5. Calibration check completed. Test results 0. _____ AC.
- (X) 6. Air blank completed.
- (X) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

DPS Form Exh G-2 (Rev 05-01)

~~#~~ Interferent detected

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

02/10/2016

16:03:14

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. JACOBS #CV23

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	16:03:47
0.100 Cal Check	INT*	16:04:08
Air Blank	0.000	16:04:39

*Interferent Detect

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 2-10-16 TIME 1105
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO-DUI Trailer

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 ~~0.100~~ ^{0.100} ~~0.100~~ ^{0.100} AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(✓) 4. Air blank completed.
(✓) 5. Calibration check completed. Test results 0.099 AC.
(✓) 6. Air blank completed.
(✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature]

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Office
DATE 2-10-14 TIME 1409
INTOXILYZER SERIAL # 80-001096 LOCATION CJMO-DUI Trailer

() 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- () 1. Clock time check.
() 2. Date check.

OPERATIONAL TESTS

- () 1. Alcohol-free subject test result 0. 000 AC.
() 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
() 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
() 4. Standard Calibration Check standard 0. 100 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

02/10/2016

16:09:54

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 02/10/1960

Sex: F Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:11:19
Diagnostic Test	Pass	16:11:53
Air Blank	0.000	16:12:21
0.100 Cal Check	0.100	16:12:42
Air Blank	0.000	16:13:11
Subject Test	0.000	16:13:37
Air Blank	ABT*	16:13:43
Air Blank	IPS**	16:13:51

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

02/10/2016

16:14:43

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 02/10/1960

Sex: F

Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	16:15:53
Diagnostic Test	Pass	16:16:28
Air Blank	0.000	16:16:55
0.100 Cal Check	0.099	16:17:16
Air Blank	0.000	16:17:46
Subject Test	0.254*	16:20:53
Air Blank	0.000	16:21:26
Air Blank	0.000	16:21:53
Subject Test	RGE**	16:22:36
Air Blank	RFI***	16:22:43

*Deficient Sample

**Range Exceeded

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Office
DATE 2-10-16 TIME 1623
INTOXILYZER SERIAL # 80001096 LOCATION CJMO - DUI Trailer

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution _____ AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. _____ AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

DPS Form Exh G-2 (Rev 05-01)

Ambient Fail - mouthwash from G-4 was
nearby

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

02/10/2016

16:23:37

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	AMB*	16:24:08
Air Blank	0.000	16:24:36

*Ambient Fail

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshals ofc
DATE 2-10-16 TIME 1624
INTOXILYZER SERIAL # 80-001096 LOCATION CUMO-DUI Trailer

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

02/10/2016

16:24:40

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	16:25:12
0.100 Cal Check	0.099	16:25:33
Air Blank	0.000	16:26:03

Preventative Maintenance
Performed -- Timer Reset