

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 3-11-16 TIME 0712
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO-Booking Ofc

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.100 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

03/11/2016

07:14:24

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	07:17:42
0.100 Cal Check	0.100	07:18:02
Air Blank	0.000	07:18:32

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 3-11-16 TIME 0715
INTOXILYZER SERIAL # 80-001096 LOCATION WMO-Booking Ofc

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
 2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0.000 AC.
 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
 4. Standard Calibration Check standard 0.100 AC. Result 0.098 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

Ambient Fail during calibration check. @0000
Passed on second try

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

03/11/2016

07:19:19

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Subject: TEST,TEST

DOB: 09/30/1963

Sex: F Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:20:33
Diagnostic Test	Pass	07:21:06
Air Blank	0.000	07:21:34
0.100 Cal Check	0.100	07:21:55
Air Blank	0.000	07:22:25
Subject Test	0.000	07:23:01
Air Blank	ABT*	07:23:18
Air Blank	IPS**	07:23:26

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
03/11/2016 07:23:39

Standard Lot#: 33814100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Subject: TEST,TEST
DOB: 09/30/1963
Sex: F Weight: 190

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	07:24:51
Diagnostic Test	Pass	07:25:25
Air Blank	0.000	07:25:52
0.100 Cal Check	0.099	07:26:13
Air Blank	0.000	07:26:43
Subject Test	0.236*	07:29:50
Air Blank	0.000	07:30:22
Air Blank	0.000	07:30:50
Subject Test	RGE**	07:32:16
Air Blank	RFI***	07:32:25

*Deficient Sample

**Range Exceeded

***RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY

STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 3-11-10 TIME 0733
INTOXILYZER SERIAL # 80-0010 LOCATION WMO-Booking Ofc

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.098 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

03/11/2016

07:33:32

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test	g/210L	Time
Air Blank	AMB*	07:34:04
Air Blank	0.000	07:34:32

*Ambient Fail

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
03/11/2016 07:34:38

Standard Lot#: 33814100A1
Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53
CAMP VERDE MO

Operator: D. HUGHES #53
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	07:35:08
0.100 Cal Check	0.098	07:35:29
Air Blank	0.000	07:35:58

Preventative Maintenance
Performed -- Timer Reset