

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Debbie Hughes AGENCY Camp Verde Marshal's Ofc
DATE 9-22-16 TIME 0831
INTOXILYZER SERIAL # 80-000096 LOCATION CYMO-Booking

- (✓) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 2.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (✓) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- (✓) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- (✓) 4. Air blank completed.
- (✓) 5. Calibration check completed. Test results 0.099 AC.
- (✓) 6. Air blank completed.
- (✓) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] #53

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

09/22/2016

08:34:40

Standard Lot#: 33814100A1

Last Changed By: D. HUGHES #53

QAS: D. HUGHES #53

CAMP VERDE MO

Operator: D. HUGHES #53

CAMP VERDE MO

Test g/210L Time

Air Blank 0.000 08:35:10
0.100 Cal Check 0.099 08:35:31
Air Blank 0.000 08:36:01

Preventative Maintenance
Performed -- Timer Reset