

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST D. Jacobs AGENCY Camp Verde Marshals
DATE 12/8/17 TIME 2245
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Badging Room

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.100 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  Sgt. C108

DPS Form Exh G-2 (Rev 05-01)

* Note: 30 Day completed upon return from Maintenance
see attached form.

INTOXILYZER 8000

Location: CAMP VERDE MO

Serial Number: 80-001096

Core Version: 8105.48

12/08/2017

22:46:10

Standard Lot#: 32316100A1

Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08

CAMP VERDE MO

Operator: D. JACOBS #CV08

CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	22:48:11
0.100 Cal Check	0.100	22:48:32
Air Blank	0.000	22:49:01

Preventative Maintenance
Performed -- Timer Reset



ARIZONA DEPARTMENT OF PUBLIC SAFETY

INTOXILYZER MAINTENANCE REPORT

CR NO. I800010960005

Page 1 of 1

AGENCY Camp Verde Marshal's Office
 Camp Verde, AZ 86322
OFFICER JACOBS,#CV08
DATE December 05, 2017



INSTRUMENT

80001096

INSTRUMENT LOCATION

Camp Verde Marshal's Office

MAINTENANCE REQUESTED

Preventative Maintenance

MAINTENANCE PERFORMED

Preventative Maintenance Performed:

Inspection, Calibration, Verification

A handwritten signature in black ink, appearing to read "Joseph M Slowinski".

JOSEPH M SLOWINSKI, #6101, Electronics Technician
Central Regional Crime Laboratory
2102 W. Encanto Blvd., Phoenix, AZ 85009
602-223-2159