

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Sgt. D. Jacobs AGENCY Camp Verde Marshals
DATE 6/24/18 TIME 1135
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Booking

- () 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.109 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- () 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- () 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- () 4. Air blank completed.
- () 5. Calibration check completed. Test results 0.099 AC.
- () 6. Air blank completed.
- () 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV08

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
06/24/2018 11:37:25

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	11:37:56
0.100 Cal Check	0.099	11:38:16
Air Blank	0.000	11:38:46

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Sgt. D. Jacobs AGENCY Camp Verde Marshals
DATE 6/24/18 TIME 1139
INTOXILYZER SERIAL # 80-001096 LOCATION CVMO Booking

(X) 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- (A) 1. Clock time check.
(X) 2. Date check.

OPERATIONAL TESTS

- (X) 1. Alcohol-free subject test result 0. 000 AC.
(X) 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
(X) 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
(X) 4. Standard Calibration Check standard 0. 098 AC. Result 0. 100 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS

SIGNATURE  CV08

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
06/24/2018

11:39:15

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/30/1989
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:40:23
Diagnostic Test	Pass	11:40:57
Air Blank	0.000	11:41:24
0.100 Cal Check	0.099	11:41:45
Air Blank	0.000	11:42:15
Subject Test	0.000	11:43:04
Air Blank	ABT*	11:43:13
Air Blank	IPS**	11:43:23

*Sequence Aborted

**Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
06/24/2018

11:43:41

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/30/1989
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:44:40
Diagnostic Test	Pass	11:45:14
Air Blank	0.000	11:45:41
0.100 Cal Check	0.098	11:46:02
Air Blank	0.000	11:46:32
Subject Test	0.040*	11:49:39
Air Blank	0.000	11:50:11
Air Blank	0.000	11:50:38
Subject Test	DEF**	11:53:46
Air Blank	0.000	11:54:14

*Deficient Sample
**Deficient Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
06/24/2018

11:54:44

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Subject: TEST,TEST
DOB: 05/30/1989
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	11:55:39
Diagnostic Test	Pass	11:56:13
Air Blank	0.000	11:56:40
0.100 Cal Check	0.098	11:57:01
Air Blank	0.000	11:57:31
Subject Test	XXX*	11:58:03
Air Blank	RFI**	11:58:21

*Invalid Sample
Begin new deprivation period
**RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Sgt. D. Jacobs AGENCY Camp Verde Marshals
DATE 6/24/18 TIME 1200
INTOXILYZER SERIAL # 8a-001096 LOCATION CVMO Booking

- (X) 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 0.100 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- (+) 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
(+) 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
(+) 4. Air blank completed.
(X) 5. Calibration check completed. Test results 0. 097 AC.
(+) 6. Air blank completed.
(+) 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE  CV08

INTOXILYZER 8000
Location: CAMP VERDE MO
Serial Number: 80-001096
Core Version: 8105.48
06/24/2018 12:00:16

Standard Lot#: 32316100A1
Last Changed By: D. JACOBS #CV08

QAS: D. JACOBS #CV08
CAMP VERDE MO

Operator: D. JACOBS #CV08
CAMP VERDE MO

Test	g/210L	Time
Air Blank	0.000	12:00:46
0.100 Cal Check	0.097	12:01:07
Air Blank	0.000	12:01:36

Preventative Maintenance
Performed -- Timer Reset