

EXHIBIT G-2  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 12-12-08 TIME 2001

INTOXILYZER SERIAL # 80-001222 LOCATION VERDE JAIL

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .100 AC.  
OR  
Pour a standard alcohol concentration solution \_\_\_\_\_ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach  $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- DM* 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- DM* 4. Air blank completed.
- DM* 5. Calibration check completed. Test results 0. 101 AC.
- DM* 6. Air blank completed.
- DM* 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE *Daniel McGinnis* 086/QAS # 33426

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

12/12/2008

20:01:38


-----  
Standard Lot#: 804301I

Last Changed By: D. RAISS #3693

-----  
QAS: D. MCGINNIS #8086

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:02:05
Cal Check	0.101	20:02:24
Air Blank	0.000	20:02:54

  
Operator's Signature

INTOXILYZER 8000  
Location: YCSO-VERDE JAIL  
Serial Number : 80-001222  
12/12/2008

20:06:19

---

Standard Lot#: 804301I  
Last Changed By: D. RAISS #3693

---

QAS: D. MCGINNIS #8086  
YAVAPAI CSO

---

Operator: D. MCGINNIS #8086  
YAVAPAI CSO

---

Subject: SQAP,TEST  
DOB: 09/09/1963  
Sex: M Weight: 222

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:07:21
Diagnostic Test	Pass	20:07:49
Air Blank	0.000	20:08:16
0.100 Cal Check	0.100	20:08:35
Air Blank	0.000	20:09:04
Subject Test	0.000	20:09:24
Air Blank	ABT*	20:09:42
Air Blank	RFI**	20:09:49

\*Sequence Aborted

\*\*RFI Detect

**Not a Successfully Completed Test Sequence**

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number : 80-001222

12/12/2008

20:10:31

---

Standard Lot#: 804301I

Last Changed By: D. RAISS

#3693

---

QAS: D. MCGINNIS

#8086

YAVAPAI CSO

---

Operator: D. MCGINNIS

#8086

YAVAPAI CSO

---

Subject: SQAP, TEST

DOB: 09/09/1963

Sex: M

Weight: 222

---

15 Minute Deprivation Period? Yes

---

Test	g/210L	Time
Air Blank	0.000	20:11:30
Diagnostic Test	Pass	20:11:57
Air Blank	0.000	20:12:24
0.100 Cal Check	0.101	20:12:43
Air Blank	0.000	20:13:13
Subject Test	0.054*	20:16:21
Air Blank	0.000	20:16:53
Air Blank	0.000	20:17:21
Subject Test	XXX**	20:18:01
Air Blank	IPS***	20:18:10

\*Deficient Sample

\*\*Invalid Sample

Begin new deprivation period

\*\*\*Improper Sample

**Not a Successfully Completed Test Sequence**



INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

12/12/2008

20:21:49

-----  
Standard Lot#: 804301I

Last Changed By: D. RAISS #3693

-----  
QAS: D. MCGINNIS #8086

YAVAPAI CSO  
-----

Test	g/210L	Time
Air Blank	0.000	20:22:16
Cal Check	0.101	20:22:35
Air Blank	0.000	20:23:05

  
Operator's Signature

EXHIBIT G-4  
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY  
STANDARD QUALITY ASSURANCE PROCEDURES  
INTOXILYZER MODEL 8000  
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO  
DATE 12-12-08 TIME 2006  
INTOXILYZER SERIAL # 80-001232 LOCATION VERDE JAIL

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- DM 1. Clock time check.  
DM 2. Date check.

OPERATIONAL TESTS

- DM 1. Alcohol-free subject test result 0. 000 AC.  
DM 2. Error recognition logic system functioning.  
Not a Successfully Completed Test Sequence printed  
DM 3. Proper sample recognition system.  
Not a Successfully Completed Test Sequence printed  
Deficient sample printed.  
DM 4. Standard Calibration Check standard 0. 100 AC. Result 0. 101 AC.

Instrument is operating properly and accurately. YES DM NO     

COMMENTS

AMBIENT FAIL ON LAST CALCHECK  
ALL IN WORKING ORDER

SIGNATURE Daniel McGinnis 8064 QAS#33426