

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 06-03-09 TIME 2349

INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR

~~Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$.~~

2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 10 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McGinnis 543/QAS#33426

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

06/03/2009

23:49:25

Standard Lot#: 901401I

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	23:49:53
Cal Check	0.100	23:50:12
Air Blank	0.000	23:50:41



Operator's Signature

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO
DATE 06-03-09 TIME 2359
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

1. Clock time check.
2. Date check.

OPERATIONAL TESTS

1. Alcohol-free subject test result 0. 000 AC. ~~100~~
2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed ~~Success~~
3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed ~~Success~~
Deficient sample printed.
4. Standard Calibration Check standard 0. 10 AC. Result 0. 102 AC.

Instrument is operating properly and accurately YES NO

COMMENTS

RFI detected during 'proper samples' due
to cell phone

All OK

SIGNATURE

Daniel McGinnis 543/QAS # 53426

DPS Form Exh G-4 (Rev 05-01)

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number : 80-001222

06/03/2009

23:51:32

Standard Lot#: 901401I

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP, TEST

DOB: 09/09/1990

Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:52:31
Diagnostic Test	Pass	23:52:58
Air Blank	0.000	23:53:25
0.100 Cal Check	0.100	23:53:44
Air Blank	0.000	23:54:14
Subject Test	0.000	23:54:55
Air Blank	ABT*	23:55:02
Air Blank	RFI**	23:55:10

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number : 80-001222
06/03/2009

23:59:24

Standard Lot#: 901401I
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:00:27
Diagnostic Test	Pass	00:00:54
Air Blank	0.000	00:01:22
0.100 Cal Check	0.101	00:01:41
Air Blank	0.000	00:02:10
Subject Test	0.046*	00:05:19
Air Blank	0.000	00:05:50
Air Blank	0.000	00:06:18
Subject Test	XXX**	00:06:40
Air Blank	IPS***	00:06:48

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number : 80-001222
06/03/2009

23:55:20

Standard Lot#: 901401I
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	23:56:19
Diagnostic Test	Pass	23:56:46
Air Blank	0.000	23:57:14
0.100 Cal Check	0.101	23:57:32
Air Blank	0.000	23:58:02
Subject Test	RFI*	23:58:44
Air Blank	0.000	23:59:17

*RFI Detect

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 06-04-09 TIME 0007

INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 102 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McGinnis 343/CAS# 33426

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

06/04/2009

00:07:49


Standard Lot#: 901401I

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	00:08:17
Cal Check	0.102	00:08:36
Air Blank	0.000	00:09:05


Operator's Signature