

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (1)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McInnis AGENCY YCSO
DATE 11-18-10 TIME 1711
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution, 10 AC OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$.
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 099 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McInnis SCS/QA # 33426

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

11/18/2010

17:11:19

Standard Lot#: 712322

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Operator: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	17:11:55
0.100 Cal Check	0.099	17:12:16
Air Blank	0.000	17:12:46

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McCann AGENCY YCSO
DATE 11-18-10^{pm} TIME 1714
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
DM 1. Clock time check
DM 2. Date check.

OPERATIONAL TESTS
DM 1. Alcohol-free subject test result 0. 00 AC.
DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
DM 4. Standard Calibration Check standard 0. 10 AC. Result 0. 099 AC.

Instrument is operating properly and accurately. YES NO

COMMENTS OK

SIGNATURE Daniel McCann S43/QAS# 33426

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
11/18/2010 17:14:44

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:15:56
Diagnostic Test	Pass	17:16:31
Air Blank	0.000	17:16:58
0.100 Cal Check	0.100	17:17:19
Air Blank	0.000	17:17:49
Subject Test	0.000	17:18:06
Air Blank	ABT*	17:18:14
Air Blank	0.000	17:18:43

*Sequence Aborted

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
11/18/2010

17:19:07

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	17:20:02
Diagnostic Test	Pass	17:20:37
Air Blank	0.000	17:21:05
0.100 Cal Check	0.099	17:21:25
Air Blank	0.000	17:21:55
Subject Test	0.106*	17:25:03
Air Blank	0.000	17:25:35
Air Blank	0.000	17:26:02
Subject Test	XXX**	17:26:38
Air Blank	IPS***	17:26:47

*Deficient Sample

**Invalid Sample

Begin new deprivation period

***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY VC50
DATE 11-18-10 TIME 1728
INTOXILYZER SERIAL # 80-80122 LOCATION Verde, Trail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR
Pour a standard alcohol concentration solution AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0.09 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McGinnis 543/QAS#38426

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

11/18/2010

17:28:38

Standard Lot#: 712322

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Operator: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	17:29:09
0.100 Cal Check	0.099	17:29:30
Air Blank	0.000	17:30:00

Preventative Maintenance
Performed -- Timer Reset

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

11/18/2010

17:27:30

Standard Lot#: 712322

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Operator: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	AMB*	17:28:02
Air Blank	0.000	17:28:30

*Ambient Fail