

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (1)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McGinnis AGENCY YCSO

DATE 12-19-10 TIME 0107

INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

- 1 Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution 10 AC
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. ~~Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach 34°C ± 0.2°C~~
- 2 Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
- 3 Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
- 4 Air blank completed.
- 5 Calibration check completed. Test results 0. 097 AC.
- 6 Air blank completed.
- 7 Remove printed record. Attach the record to the completed checklist.

SIGNATURE

Daniel McGinnis 343 / QAS # 33420

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
12/19/2010 01:07:03

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	01:07:56
0.100 Cal Check	0.097	01:08:17
Air Blank	0.000	01:08:47

Preventative Maintenance
Performed -- Timer Reset