

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McInnis AGENCY YCSO
DATE 2-8-11 TIME 2207
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

- DM* 1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR
Pour a standard alcohol concentration solution ~~_____~~ AC into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
- DM* 2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
DM 3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
DM 4. Air blank completed.
DM 5. Calibration check completed. Test results 0. 100 AC.
DM 6. Air blank completed.
DM 7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE Daniel McInnis 543/QAS# 33426

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel McQuinn AGENCY YCSO
DATE 2-8-11 TIME 2203
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS
DM 1. Clock time check.
DM 2. Date check.

OPERATIONAL TESTS
DM 1. Alcohol-free subject test result 0. 00 AC.
DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
DM 4. Standard Calibration Check standard 0. 10 AC. Result 0. 10 AC.

Instrument is operating properly and accurately. YES DM NO _____

COMMENTS All OK

SIGNATURE Daniel McQuinn 543/QAS #53426

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
02/08/2011

22:03:55

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:04:51
Diagnostic Test	Pass	22:05:25
Air Blank	0.000	22:05:53
0.100 Cal Check	0.100	22:06:14
Air Blank	0.000	22:06:43
Subject Test	0.000	22:07:01
Air Blank	ABT*	22:07:08
Air Blank	RFI**	22:07:16

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
02/08/2011

22:07:28

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP,TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	22:08:27
Diagnostic Test	Pass	22:09:01
Air Blank	0.000	22:09:28
0.100 Cal Check	0.101	22:09:49
Air Blank	0.000	22:10:19
Subject Test	0.073*	22:13:26
Air Blank	0.000	22:14:00
Air Blank	0.000	22:14:28
Subject Test	XXX**	22:14:41
Air Blank	IPS***	22:14:48

*Deficient Sample
**Invalid Sample
Begin new deprivation period
***Improper Sample

Not a Successfully Completed Test Sequence

INTOXILYZER 8000

Location: YCSO-VERDE JAIL

Serial Number: 80-001222

Core Version: 8105.48

02/08/2011

22:16:20

Standard Lot#: 712322

Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086

YAVAPAI CSO

Operator: D. MCGINNIS #8086

YAVAPAI CSO

Test	g/210L	Time
Air Blank	AMB*	22:16:51
Air Blank	0.000	22:17:19

*Ambient Fail

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
02/08/2011 22:17:22

Standard Lot#: 712322
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	22:17:53
0.100 Cal Check	0.100	22:18:14
Air Blank	0.000	22:18:44

Preventative Maintenance
Performed -- Timer Reset