

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
04/04/2011

23:59:48

Standard Lot#: 35710100A2
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test	g/210L	Time
Air Blank	0.000	00:00:19
0.100 Cal Check	0.099	00:00:40
Air Blank	0.000	00:01:10

Preventative Maintenance
Performed -- Timer Reset

EXHIBIT G-4
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000

STANDARD QUALITY ASSURANCE PROCEDURE

QA SPECIALIST Daniel M'Ginnis AGENCY YCSO
DATE 04-05-11 TIME 0007
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Trail

DM 1. Display Reads "PUSH BUTTON TO START"

DIAGNOSTIC TESTS

- DM 1. Clock time check.
- DM 2. Date check.

OPERATIONAL TESTS

- DM 1. Alcohol-free subject test result 0, 000 AC.
- DM 2. Error recognition logic system functioning.
Not a Successfully Completed Test Sequence printed
- DM 3. Proper sample recognition system.
Not a Successfully Completed Test Sequence printed
Deficient sample printed.
- DM 4. Standard Calibration Check standard 0, 10 AC. Result 0, 098 AC.

Instrument is operating properly and accurately. YES DM NO

COMMENTS All ok

SIGNATURE Daniel M'Ginnis 543/QAS# 53426

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
04/05/2011

00:02:10

Standard Lot#: 35710100A2
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:04:13
Diagnostic Test	Pass	00:04:47
Air Blank	0.000	00:05:14
0.100 Cal Check	0.098	00:05:35
Air Blank	0.000	00:06:05
Subject Test	0.000	00:06:30
Air Blank	ABT*	00:06:38
Air Blank	RFI**	00:06:47

*Sequence Aborted

**RFI Detect

Not a Successfully Completed Test Sequence

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
04/05/2011

00:07:18

Standard Lot#: 35710100A2
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Subject: SQAP, TEST
DOB: 09/09/1990
Sex: M Weight: 200

15 Minute Deprivation Period? Yes

Test	g/210L	Time
Air Blank	0.000	00:08:14
Diagnostic Test	Pass	00:08:48
Air Blank	0.000	00:09:15
0.100 Cal Check	0.099	00:09:36
Air Blank	0.000	00:10:06
Subject Test	0.059*	00:13:13
Air Blank	0.000	00:13:46
Air Blank	0.000	00:14:13
Subject Test	XXX**	00:14:41
Air Blank	IPS***	00:14:49

*Deficient Sample
**Invalid Sample
 Begin new deprivation period
***Improper Sample

Not a Successfully Completed Test Sequence

EXHIBIT G-2
THIS REPORT PREPARED PURSUANT TO DUTY IMPOSED BY A.A.C. R13-10-104 (A)

ARIZONA DEPARTMENT OF PUBLIC SAFETY
STANDARD QUALITY ASSURANCE PROCEDURES
INTOXILYZER MODEL 8000
STANDARD CALIBRATION CHECK PROCEDURE

QA SPECIALIST Daniel McConnis AGENCY YCSO
DATE 04-05-11 TIME 0017
INTOXILYZER SERIAL # 80-001222 LOCATION Verde Jail

1. Ensure that gas tank is attached to instrument and contains a standard alcohol concentration solution .10 AC.
OR
Pour a standard alcohol concentration solution _____ AC, into a clean dry simulator and assemble the simulator. Ensure that a tight seal has been made. Turn on the simulator and allow temperature to reach $34^{\circ}\text{C} \pm 0.2^{\circ}\text{C}$
2. Intoxilyzer 8000 display reads "PUSH BUTTON TO START"
3. Go to the "Control Testing Menu". Select "D" for dry control test or "W" for wet control test. After selection is made press ENTER.
4. Air blank completed.
5. Calibration check completed. Test results 0. 0.05 AC.
6. Air blank completed.
7. Remove printed record. Attach the record to the completed checklist.

SIGNATURE [Signature] 543/QAs # 33424

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
04/05/2011 00:16:45

Standard Lot#: 35710100A2
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test	g/210L	Time
Air Blank	AMB*	00:17:17
Air Blank	0.000	00:17:46

*Ambient Fail

INTOXILYZER 8000
Location: YCSO-VERDE JAIL
Serial Number: 80-001222
Core Version: 8105.48
04/05/2011 00:17:51

Standard Lot#: 35710100A2
Last Changed By: D. MCGINNIS #8086

QAS: D. MCGINNIS #8086
YAVAPAI CSO

Operator: D. MCGINNIS #8086
YAVAPAI CSO

Test g/210L Time

Air Blank 0.000 00:18:22
0.100 Cal Check 0.098 00:18:43
Air Blank 0.000 00:19:13

Preventative Maintenance
Performed -- Timer Reset